

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100324\
 Data File : VX043266.D
 Acq On : 03 Oct 2024 16:45
 Operator : JC/MD
 Sample : P4263-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB04-20241001

Quant Time: Oct 04 01:34:48 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	77122	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	145209	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	130345	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	54506	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	65028	51.629	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 103.260%			
35) Dibromofluoromethane	5.385	113	47859	47.791	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 95.580%			
50) Toluene-d8	8.647	98	173702	50.053	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 100.100%			
62) 4-Bromofluorobenzene	11.079	95	62736	49.760	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 99.520%			
Target Compounds						
16) Acetone	2.386	43	1697	3.226	ug/l #	86
30) Chloroform	5.093	83	34658	19.132	ug/l	95
47) Bromodichloromethane	7.824	83	3119	2.103	ug/l #	94
52) Toluene	8.720	92	3925	1.633	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100324\
Data File : VX043266.D
Acq On : 03 Oct 2024 16:45
Operator : JC/MD
Sample : P4263-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB04-20241001

Quant Time: Oct 04 01:34:48 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 16:50:57 2024
Response via : Initial Calibration

