

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100324\
 Data File : VX043267.D
 Acq On : 03 Oct 2024 17:08
 Operator : JC/MD
 Sample : P4263-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE115D2-20240930

Quant Time: Oct 04 01:35:14 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	82097	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	158670	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	139467	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	60828	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	70673	52.711	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 105.420%			
35) Dibromofluoromethane	5.385	113	51568	47.126	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 94.260%			
50) Toluene-d8	8.647	98	182486	48.123	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 96.240%			
62) 4-Bromofluorobenzene	11.079	95	67488	48.988	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 97.980%			
Target Compounds					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.319	101	56794	61.831	ug/l	99
12) 1,1-Dichloroethene	2.312	96	33975	37.248	ug/l	98
16) Acetone	2.392	43	974	1.740	ug/l	92
24) 1,1-Dichloroethane	3.611	63	8697	4.953	ug/l	98
27) cis-1,2-Dichloroethene	4.483	96	5263	4.649	ug/l	88
30) Chloroform	5.092	83	3135	1.626	ug/l	89
32) 1,1,1-Trichloroethane	5.385	97	4110	2.320	ug/l	97
38) Carbon Tetrachloride	5.672	117	4139	2.529	ug/l	90
44) Trichloroethene	7.123	130	1388493	1249.010	ug/l	95
49) 1,4-Dioxane	7.683	88	717	27.416	ug/l	97
55) 1,1,2-Trichloroethane	9.159	97	2228	2.134	ug/l	89
64) Tetrachloroethene	9.275	164	1121	1.158	ug/l	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100324\
Data File : VX043267.D
Acq On : 03 Oct 2024 17:08
Operator : JC/MD
Sample : P4263-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE115D2-20240930

Quant Time: Oct 04 01:35:14 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 16:50:57 2024
Response via : Initial Calibration

