

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100324\
 Data File : VX043270.D
 Acq On : 03 Oct 2024 18:17
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 04 01:36:46 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	105	0.00
2 T	Dichlorodifluoromethane	50.000	46.016	8.0	94	0.00
3 P	Chloromethane	50.000	42.152	15.7	88	0.00
4 C	Vinyl Chloride	50.000	45.656	8.7#	95	0.00
5 T	Bromomethane	50.000	39.901	20.2	86	0.00
6 T	Chloroethane	50.000	43.600	12.8	91	0.00
7 T	Trichlorofluoromethane	50.000	45.551	8.9	99	0.00
8 T	Diethyl Ether	50.000	46.990	6.0	100	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	47.024	6.0	99	0.00
10 T	Methyl Iodide	50.000	43.769	12.5	90	0.00
11 T	Tert butyl alcohol	250.000	246.020	1.6	101	0.00
12 CM	1,1-Dichloroethene	50.000	44.843	10.3#	95	0.00
13 T	Acrolein	250.000	232.213	7.1	97	0.00
14 T	Allyl chloride	50.000	44.978	10.0	94	0.00
15 T	Acrylonitrile	250.000	236.653	5.3	99	0.00
16 T	Acetone	250.000	205.382	17.8	93	0.00
17 T	Carbon Disulfide	50.000	40.390	19.2	87	0.00
18 T	Methyl Acetate	50.000	50.357	-0.7	102	0.00
19 T	Methyl tert-butyl Ether	50.000	49.884	0.2	104	0.00
20 T	Methylene Chloride	50.000	44.579	10.8	96	0.00
21 T	trans-1,2-Dichloroethene	50.000	47.462	5.1	100	0.00
22 T	Diisopropyl ether	50.000	48.370	3.3	101	0.00
23 T	Vinyl Acetate	250.000	252.310	-0.9	100	0.00
24 P	1,1-Dichloroethane	50.000	48.053	3.9	101	0.00
25 T	2-Butanone	250.000	230.338	7.9	96	0.00
26 T	2,2-Dichloropropane	50.000	45.171	9.7	96	0.00
27 T	cis-1,2-Dichloroethene	50.000	46.137	7.7	98	0.00
28 T	Bromochloromethane	50.000	41.623	16.8	91	0.00
29 T	Tetrahydrofuran	250.000	233.566	6.6	100	0.00
30 C	Chloroform	50.000	47.125	5.8#	101	0.00
31 T	Cyclohexane	50.000	46.249	7.5	99	0.00
32 T	1,1,1-Trichloroethane	50.000	47.967	4.1	100	0.00
33 S	1,2-Dichloroethane-d4	50.000	42.375	15.3	89	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	100	0.00
35 S	Dibromofluoromethane	50.000	44.914	10.2	91	0.00
36 T	1,1-Dichloropropene	50.000	48.605	2.8	98	0.00
37 T	Ethyl Acetate	50.000	49.987	0.0	101	0.00
38 T	Carbon Tetrachloride	50.000	50.483	-1.0	99	0.00
39 T	Methylcyclohexane	50.000	48.692	2.6	97	0.00
40 TM	Benzene	50.000	49.037	1.9	99	0.00
41 T	Methacrylonitrile	50.000	50.698	-1.4	96	0.00
42 TM	1,2-Dichloroethane	50.000	50.714	-1.4	100	0.00
43 T	Isopropyl Acetate	50.000	49.663	0.7	97	0.00
44 TM	Trichloroethene	50.000	49.636	0.7	102	0.00
45 C	1,2-Dichloropropane	50.000	49.467	1.1#	98	0.00
46 T	Dibromomethane	50.000	49.900	0.2	99	0.00
47 T	Bromodichloromethane	50.000	50.614	-1.2	98	0.00
48 T	Methyl methacrylate	50.000	49.253	1.5	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX100324\
 Data File : VX043270.D
 Acq On : 03 Oct 2024 18:17
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 04 01:36:46 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1101.934	-10.2	107	0.00
50 S	Toluene-d8	50.000	43.208	13.6	85	0.00
51 T	4-Methyl-2-Pentanone	250.000	248.328	0.7	96	0.00
52 CM	Toluene	50.000	48.458	3.1#	95	0.00
53 T	t-1,3-Dichloropropene	50.000	49.581	0.8	93	0.00
54 T	cis-1,3-Dichloropropene	50.000	49.119	1.8	96	0.00
55 T	1,1,2-Trichloroethane	50.000	49.378	1.2	96	0.00
56 T	Ethyl methacrylate	50.000	51.068	-2.1	95	0.00
57 T	1,3-Dichloropropane	50.000	49.264	1.5	96	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	255.903	-2.4	94	0.00
59 T	2-Hexanone	250.000	250.379	-0.2	96	0.00
60 T	Dibromochloromethane	50.000	48.262	3.5	91	0.00
61 T	1,2-Dibromoethane	50.000	48.430	3.1	94	0.00
62 S	4-Bromofluorobenzene	50.000	43.222	13.6	83	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	95	0.00
64 T	Tetrachloroethene	50.000	48.611	2.8	94	0.00
65 PM	Chlorobenzene	50.000	48.105	3.8	93	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	50.111	-0.2	93	0.00
67 C	Ethyl Benzene	50.000	49.007	2.0#	92	0.00
68 T	m/p-Xylenes	100.000	96.606	3.4	92	0.00
69 T	o-Xylene	50.000	48.428	3.1	93	0.00
70 T	Styrene	50.000	50.463	-0.9	94	0.00
71 P	Bromoform	50.000	50.991	-2.0	93	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	97	0.00
73 T	Isopropylbenzene	50.000	47.711	4.6	92	0.00
74 T	N-amyl acetate	50.000	49.107	1.8	96	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	49.035	1.9	97	0.00
76 T	1,2,3-Trichloropropane	50.000	47.996	4.0	97	0.00
77 T	Bromobenzene	50.000	47.537	4.9	94	0.00
78 T	n-propylbenzene	50.000	49.068	1.9	93	0.00
79 T	2-Chlorotoluene	50.000	47.065	5.9	94	0.00
80 T	1,3,5-Trimethylbenzene	50.000	48.704	2.6	95	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	45.280	9.4	90	0.00
82 T	4-Chlorotoluene	50.000	48.101	3.8	95	0.00
83 T	tert-Butylbenzene	50.000	47.868	4.3	94	0.00
84 T	1,2,4-Trimethylbenzene	50.000	48.122	3.8	94	0.00
85 T	sec-Butylbenzene	50.000	48.800	2.4	94	0.00
86 T	p-Isopropyltoluene	50.000	48.657	2.7	94	0.00
87 T	1,3-Dichlorobenzene	50.000	48.287	3.4	96	0.00
88 T	1,4-Dichlorobenzene	50.000	46.577	6.8	95	0.00
89 T	n-Butylbenzene	50.000	50.016	-0.0	94	0.00
90 T	Hexachloroethane	50.000	48.507	3.0	92	0.00
91 T	1,2-Dichlorobenzene	50.000	47.691	4.6	96	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	47.920	4.2	93	0.00
93 T	1,2,4-Trichlorobenzene	50.000	49.578	0.8	95	0.00
94 T	Hexachlorobutadiene	50.000	50.601	-1.2	96	0.00
95 T	Naphthalene	50.000	48.276	3.4	95	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100324\
Data File : VX043270.D
Acq On : 03 Oct 2024 18:17
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Oct 04 01:36:46 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 16:50:57 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	48.189	3.6	94	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6