

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100325\
 Data File : VX047983.D
 Acq On : 03 Oct 2025 10:16
 Operator : JC/MD
 Sample : VX1003WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1003WBL01

Quant Time: Oct 03 23:42:47 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	132031	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	278145	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	290715	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	147497	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	118862	56.558	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	113.120%
35) Dibromofluoromethane	5.373	113	95126	50.995	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.000%
50) Toluene-d8	8.634	98	318678	49.372	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.740%
62) 4-Bromofluorobenzene	11.067	95	139510	56.951	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	113.900%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100325\
Data File : VX047983.D
Acq On : 03 Oct 2025 10:16
Operator : JC/MD
Sample : VX1003WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1003WBL01

Quant Time: Oct 03 23:42:47 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

