

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100325\
 Data File : VX047998.D
 Acq On : 03 Oct 2025 16:36
 Operator : JC/MD
 Sample : Q3248-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-184-GW-580-582

Quant Time: Oct 03 23:51:57 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	116663	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	235290	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	247287	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	124667	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	99029	53.328	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 106.660%	
35) Dibromofluoromethane	5.373	113	79228	50.208	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.420%	
50) Toluene-d8	8.635	98	267943	49.073	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 98.140%	
62) 4-Bromofluorobenzene	11.061	95	116425	56.184	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 112.360%	
Target Compounds						
						Qvalue
16) Acetone	2.380	43	14323	17.739	ug/l	93
19) Methyl tert-butyl Ether	3.111	73	25908	5.120	ug/l	97
25) 2-Butanone	4.568	43	4136	4.107	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100325\
Data File : VX047998.D
Acq On : 03 Oct 2025 16:36
Operator : JC/MD
Sample : Q3248-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-184-GW-580-582

Quant Time: Oct 03 23:51:57 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

