

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100325\
 Data File : VX048001.D
 Acq On : 03 Oct 2025 17:38
 Operator : JC/MD
 Sample : Q3248-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-184-GW-640-642

Quant Time: Oct 03 23:53:04 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	114921	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	230401	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	237241	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	118357	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	91711	50.136	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	100.280%
35) Dibromofluoromethane	5.373	113	76762	49.678	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.360%
50) Toluene-d8	8.635	98	258496	48.347	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.700%
62) 4-Bromofluorobenzene	11.067	95	108760	53.598	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	107.200%
Target Compounds						
16) Acetone	2.380	43	8407	10.570	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100325\
Data File : VX048001.D
Acq On : 03 Oct 2025 17:38
Operator : JC/MD
Sample : Q3248-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-184-GW-640-642

Quant Time: Oct 03 23:53:04 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

