

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100418\
 Data File : VX004987.D
 Acq On : 03 Oct 2018 23:21
 Operator : JC/MD
 Sample : PB113446ZHE#13
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB113446ZHE#13

Quant Time: Oct 04 01:28:42 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 03 04:00:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	290713	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	471144	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	441594	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	248106	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	201555	48.53	ug/l	0.00
Spiked Amount			Recovery	=	97.06%	
35) Dibromofluoromethane	5.51	113	177198	44.38	ug/l	0.00
Spiked Amount			Recovery	=	88.76%	
50) Toluene-d8	8.72	98	647302	48.76	ug/l	0.00
Spiked Amount			Recovery	=	97.52%	
62) 4-Bromofluorobenzene	11.14	95	241018	50.40	ug/l	0.00
Spiked Amount			Recovery	=	100.80%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
6) Chloroethane	1.73	64	154	Below Cal	#	76
11) Tert butyl alcohol	3.02	59	6078	5.234	ug/l #	79
16) Acetone	2.45	43	21875	9.022	ug/l	95
18) Methyl Acetate	2.78	43	14348	2.336	ug/l	97
20) Methylene Chloride	2.86	84	5409	0.707	ug/l	95
25) 2-Butanone	4.71	43	5457	1.527	ug/l	97
29) Tetrahydrofuran	5.17	42	2327	1.048	ug/l #	80
31) Cyclohexane	5.66	56	4766	0.902	ug/l #	1
36) 1,1-Dichloropropene	5.66	75	18734	3.439	ug/l #	46
38) Carbon Tetrachloride	5.66	117	25458	4.445	ug/l #	17
43) Isopropyl Acetate	6.46	43	247886	23.749	ug/l	95
48) Methyl methacrylate	7.88	41	22850	4.823	ug/l #	44
51) 4-Methyl-2-Pentanone	8.62	43	3291	0.501	ug/l #	37
52) Toluene	8.79	92	2007	0.214	ug/l	97
54) cis-1,3-Dichloropropene	8.62	75	1670	0.287	ug/l #	70
56) Ethyl methacrylate	9.02	69	120	2.879	ug/l #	1
58) 2-Chloroethyl Vinyl ether	8.38	63	43	14.254	ug/l #	49
59) 2-Hexanone	9.55	43	16096	2.939	ug/l #	54
68) m/p-Xylenes	10.36	106	4090	0.582	ug/l	87
69) o-Xylene	10.69	106	1417	0.220	ug/l	96
70) Styrene	10.71	104	385	2.171	ug/l #	31
74) N-amyl acetate	10.90	43	178	1.768	ug/l #	25
76) 1,2,3-Trichloropropane	11.14	75	115572	20.783	ug/l #	35
81) trans-1,4-Dichloro-2-buten	11.14	75	115572	56.417	ug/l #	10
84) 1,2,4-Trimethylbenzene	11.81	105	6985	0.276	ug/l	99
95) Naphthalene	13.83	128	2433	0.836	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.66 min Scan# 739

Delta R.T. -0.00 min

Lab File: VX004987.D

Acq: 03 Oct 2018 23:21

Instrument :

MSVOA_X

ClientSampleId :


PB113446ZHE#13

Tot Ion:168 Resp: 290713

Ion Ratio Lower Upper




168 100

99 46.4 39.3 58.9



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0



#6

Chloroethane

Concen: Below Cal

RT: 1.73 min Scan# 94

Delta R.T. 0.01 min

Lab File: VX004987.D


Acq: 03 Oct 2018 23:21

Tgt Ion: 64 Resp: 154

Ion Ratio Lower Upper



64 100


66 42.7 23.8 35.6#



Abundance Ion 63.90 (63.60 to 64.60): VX0

Ion 65.90 (65.60 to 66.60): VX0









#11

Tert butyl alcohol
 Concen: 5.234 ug/l
 RT: 3.02 min Scan# 306
 Delta R.T. -0.06 min
 Lab File: VX004987.D
 Acq: 03 Oct 2018 23:21

Instrument :
 MSVOA_X
 ClientSampleId :
 PB113446ZHE#13




Tot Ion: 59 Resp: 6078
 Ion Ratio Lower Upper
 59 100
 57 2.7 8.6 12.8#




#16

Acetone
 Concen: 9.022 ug/l
 RT: 2.45 min Scan# 212
 Delta R.T. -0.00 min
 Lab File: VX004987.D
 Acq: 03 Oct 2018 23:21

Tgt Ion: 43 Resp: 21875
 Ion Ratio Lower Upper
 43 100
 58 29.9 21.8 32.6









#18
Methvl Acetate
Concen: 2.336 ug/l
RT: 2.78 min Scan# 267
Delta R.T. -0.00 min
Lab File: VX004987.D
Acq: 03 Oct 2018 23:21




Instrument :
MSVOA_X
ClientSampleId :
PB113446ZHE#13


Tot Ion: 43 Resp: 14348
Ion Ratio Lower Upper
43 100
74 23.9 17.8 26.8



#20
Methylene Chloride
Concen: 0.707 ug/l
RT: 2.86 min Scan# 280
Delta R.T. 0.01 min
Lab File: VX004987.D
Acq: 03 Oct 2018 23:21

Tgt Ion: 84 Resp: 5409
Ion Ratio Lower Upper
84 100
49 122.6 92.3 138.5
51 37.6 31.8 47.6
86 62.3 48.5 72.7





#25

2-Butanone

Concen: 1.527 ug/l

RT: 4.71 min Scan# 583

Delta R.T. -0.00 min

Lab File: VX004987.D

Acq: 03 Oct 2018 23:21

Instrument :

MSVOA_X

ClientSampleId :





PB113446ZHE#13

Tot Ion: 43 Resp: 5457

Ion Ratio Lower Upper

43 100

72 27.5 20.8 31.2



#29

Tetrahydrofuran

Concen: 1.048 ug/l

RT: 5.17 min Scan# 659

Delta R.T. 0.01 min

Lab File: VX004987.D

Acq: 03 Oct 2018 23:21




Tgt Ion: 42 Resp: 2327


Ion Ratio Lower Upper

42 100

72 17.0 33.7 50.5#

71 40.1 32.2 48.4









#31
Cvclohexane
Concen: 0.902 ug/l
RT: 5.66 min Scan# 740
Delta R.T. 0.09 min
Lab File: VX004987.D
Acq: 03 Oct 2018 23:21

Instrument :
MSVOA_X
ClientSampleId :
PB113446ZHE#13

Tot Ion: 56 Resp: 4766
Ion Ratio Lower Upper
56 100
69 177.6 25.9 38.9#
84 145.7 71.9 107.9#




Abundance Ion 56.00 (55.70 to 56.70): VX0
Ion 69.00 (68.70 to 69.70): VX0
Ion 84.00 (83.70 to 84.70): VX0






#33
1,2-Dichloroethane-d4
Concen: 48.534 ug/l
RT: 6.06 min Scan# 805
Delta R.T. -0.01 min
Lab File: VX004987.D
Acq: 03 Oct 2018 23:21

Tgt Ion: 65 Resp: 201555
Ion Ratio Lower Upper
65 100
67 53.6 0.0 107.4



Abundance Ion 64.90 (64.60 to 65.60): VX0
Ion 66.90 (66.60 to 67.60): VX0









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.87 min Scan# 937
Delta R.T. 0.01 min
Lab File: VX004987.D
Acq: 03 Oct 2018 23:21

Instrument :
MSVOA_X
ClientSampleId :
PB113446ZHE#13

Tot Ion:114 Resp: 471144
Ion Ratio Lower Upper
114 100
63 18.9 0.0 39.8
88 15.0 0.0 27.0




Abundance Ion 113.90 (113.60 to 114.60): V
Ion 63.00 (62.70 to 63.70): VX0
Ion 87.90 (87.60 to 88.60): VX0






#35
Dibromofluoromethane
Concen: 44.377 ug/l
RT: 5.51 min Scan# 714
Delta R.T. 0.01 min
Lab File: VX004987.D
Acq: 03 Oct 2018 23:21

Tgt Ion:113 Resp: 177198
Ion Ratio Lower Upper
113 100
111 100.1 77.0 115.4
192 23.5 17.0 25.6



Abundance Ion 112.80 (112.50 to 113.50): V
Ion 110.80 (110.50 to 111.50): V
Ion 191.70 (191.40 to 192.40): V





#36

1,1-Dichloropropene

Concen: 3.439 ug/l

RT: 5.66 min Scan# 740

Delta R.T. -0.13 min

Lab File: VX004987.D

Acq: 03 Oct 2018 23:21

Instrument :

MSVOA_X

ClientSampleId :

PB113446ZHE#13


Tot Ion: 75 Resp: 18734

Ion Ratio Lower Upper

75 100

110 9.4 20.0 59.9#




77 0.0 27.1 40.7#



Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 109.90 (109.60 to 110.60): V

Ion 76.90 (76.60 to 77.60): VX0



#38

Carbon Tetrachloride

Concen: 4.445 ug/l

RT: 5.66 min Scan# 739

Delta R.T. -0.12 min

Lab File: VX004987.D

Acq: 03 Oct 2018 23:21


Tgt Ion: 117 Resp: 25458

Ion Ratio Lower Upper

117 100

119 5.1 75.2 112.8#



121 0.0 22.5 33.7#




Abundance Ion 116.80 (116.50 to 117.50): V

Ion 118.80 (118.50 to 119.50): V

Ion 120.80 (120.50 to 121.50): V





#43

Isopropyl Acetate

Concen: 23.749 ug/l

RT: 6.46 min Scan# 871

Delta R.T. -0.00 min

Lab File: VX004987.D

Acq: 03 Oct 2018 23:21

Instrument :

MSVOA_X

ClientSampleId :

PB113446ZHE#13


Tot Ion: 43 Resp: 247886

Ion Ratio Lower Upper




43 100

61 20.3 14.2 21.4

87 13.0 9.3 13.9



Abundance Ion 43.00 (42.70 to 43.70): VX0
Ion 61.00 (60.70 to 61.70): VX0
Ion 87.00 (86.70 to 87.70): VX0



#48

Methyl methacrylate

Concen: 4.823 ug/l

RT: 7.88 min Scan# 1104

Delta R.T. 0.10 min

Lab File: VX004987.D

Acq: 03 Oct 2018 23:21


Tgt Ion: 41 Resp: 22850

Ion Ratio Lower Upper



41 100


69 0.0 63.8 95.6#

39 52.5 40.7 61.1



Abundance Ion 41.00 (40.70 to 41.70): VX0
Ion 69.00 (68.70 to 69.70): VX0
Ion 39.00 (38.70 to 39.70): VX0









#50
Toluene-d8
Concen: 48.762 ug/l
RT: 8.72 min Scan# 1241
Delta R.T. -0.00 min
Lab File: VX004987.D
Acq: 03 Oct 2018 23:21




Instrument :
MSVOA_X
ClientSampleId :
PB113446ZHE#13


Tot Ion: 98 Resp: 647302
Ion Ratio Lower Upper
98 100
100 63.8 51.9 77.9



#51
4-Methyl-2-Pentanone
Concen: 0.501 ug/l
RT: 8.62 min Scan# 1225
Delta R.T. -0.03 min
Lab File: VX004987.D
Acq: 03 Oct 2018 23:21

Tgt Ion: 43 Resp: 3291
Ion Ratio Lower Upper
43 100
58 0.0 30.5 45.7#





#52

Toluene

Concen: 0.214 ug/l

RT: 8.79 min Scan# 1253

Delta R.T. 0.01 min

Lab File: VX004987.D

Acq: 03 Oct 2018 23:21

Instrument :

MSVOA_X

ClientSampleId :


PB113446ZHE#13

Tot Ion: 92 Resp: 2007

Ion Ratio Lower Upper




92 100

91 175.3 136.7 205.1



Abundance Ion 92.00 (91.70 to 92.70): VX0

Ion 91.00 (90.70 to 91.70): VX0



#54

cis-1,3-Dichloropropene

Concen: 0.287 ug/l

RT: 8.62 min Scan# 1225

Delta R.T. 0.18 min

Lab File: VX004987.D

Acq: 03 Oct 2018 23:21


Tgt Ion: 75 Resp: 1670

Ion Ratio Lower Upper

75 100

77 5.9 25.7 38.5#



39 35.8 38.3 57.5#




Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 76.90 (76.60 to 77.60): VX0

Ion 39.00 (38.70 to 39.70): VX0





#56

Ethyl methacrylate

Concen: 2.879 ug/l

RT: 9.02 min Scan# 1291

Delta R.T. -0.16 min

Lab File: VX004987.D

Acq: 03 Oct 2018 23:21

Instrument :

MSVOA_X

ClientSampleId :

PB113446ZHE#13


Tot Ion: 69 Resp: 120

Ion Ratio Lower Upper

69 100

41 665.8 56.9 85.3#


39 224.2 28.4 42.6#





Abundance Ion 69.00 (68.70 to 69.70): VX0

Ion 41.00 (40.70 to 41.70): VX0

Ion 39.00 (38.70 to 39.70): VX0



Time-->



#58

2-Chloroethyl Vinyl ether

Concen: 14.254 ug/l

RT: 8.38 min Scan# 1186

Delta R.T. 0.07 min

Lab File: VX004987.D


Acq: 03 Oct 2018 23:21

Tgt Ion: 63 Resp: 43

Ion Ratio Lower Upper


63 100

106 0.0 21.0 31.6#





Abundance Ion 62.90 (62.60 to 63.60): VX0

Ion 105.90 (105.60 to 106.60): V



Time-->









#59
2-Hexanone
Concen: 2.939 ug/l
RT: 9.55 min Scan# 1377
Delta R.T. 0.05 min
Lab File: VX004987.D
Acq: 03 Oct 2018 23:21




Instrument :
MSVOA_X
ClientSampleId :
PB113446ZHE#13


Tot Ion: 43 Resp: 16096
Ion Ratio Lower Upper
43 100
58 22.7 28.4 85.3#



#62
4-Bromofluorobenzene
Concen: 50.398 ug/l
RT: 11.14 min Scan# 1638
Delta R.T. -0.00 min
Lab File: VX004987.D
Acq: 03 Oct 2018 23:21

Tgt Ion: 95 Resp: 241018
Ion Ratio Lower Upper
95 100
174 91.0 0.0 185.0
176 88.4 0.0 180.2









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.12 min Scan# 1470
Delta R.T. -0.00 min
Lab File: VX004987.D
Acq: 03 Oct 2018 23:21

Instrument :
MSVOA_X
ClientSampleId :
PB113446ZHE#13

Tot Ion:117 Resp: 441594
Ion Ratio Lower Upper
117 100
82 54.9 42.2 63.4
119 32.4 27.4 41.0




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): VX0
Ion 118.90 (118.60 to 119.60): V






#68
m/p-Xylenes
Concen: 0.582 ug/l
RT: 10.36 min Scan# 1510
Delta R.T. -0.01 min
Lab File: VX004987.D
Acq: 03 Oct 2018 23:21

Tgt Ion:106 Resp: 4090
Ion Ratio Lower Upper
106 100
91 216.4 157.5 236.3



Abundance Ion 106.00 (105.70 to 106.70): V
Ion 91.00 (90.70 to 91.70): VX0









#69
o-Xylene
Concen: 0.220 ug/l
RT: 10.69 min Scan# 1565
Delta R.T. -0.01 min
Lab File: VX004987.D
Acq: 03 Oct 2018 23:21

Instrument :
MSVOA_X
ClientSampleId :
PB113446ZHE#13

Tot Ion:106 Resp: 1417
Ion Ratio Lower Upper
106 100
91 222.4 108.3 324.8




Abundance Ion 106.00 (105.70 to 106.70): V
Ion 91.00 (90.70 to 91.70): VX0






#70
Styrene
Concen: 2.171 ug/l
RT: 10.71 min Scan# 1567
Delta R.T. -0.01 min
Lab File: VX004987.D
Acq: 03 Oct 2018 23:21

Tgt Ion:104 Resp: 385
Ion Ratio Lower Upper
104 100
78 116.9 40.0 60.0#
103 86.5 44.3 66.5#



Abundance Ion 104.00 (103.70 to 104.70): V
Ion 78.00 (77.70 to 78.70): VX0
Ion 103.00 (102.70 to 103.70): V





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.08 min Scan# 1792

Delta R.T. -0.00 min

Lab File: VX004987.D

Acq: 03 Oct 2018 23:21

Instrument :

MSVOA_X

ClientSampleId :

PB113446ZHE#13


Tot Ion:152 Resp: 248106

Ion Ratio Lower Upper

152 100

115 55.3 40.2 120.6




150 156.1 0.0 351.0



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V



#74

N-ethyl acetate

Concen: 1.768 ug/l

RT: 10.90 min Scan# 1598

Delta R.T. -0.01 min

Lab File: VX004987.D

Acq: 03 Oct 2018 23:21

Tgt Ion: 43 Resp: 178


Ion Ratio Lower Upper

43 100

70 0.0 32.6 48.8#

55 85.4 21.6 32.4#

61 0.0 19.1 28.7#






Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 70.00 (69.70 to 70.70): VX0

Ion 55.00 (54.70 to 55.70): VX0

Ion 61.00 (60.70 to 61.70): VX0









#76
 1,2,3-Trichloropropane
 Concen: 20.783 ug/l
 RT: 11.14 min Scan# 1638
 Delta R.T. -0.16 min
 Lab File: VX004987.D
 Acq: 03 Oct 2018 23:21




Instrument :
 MSVOA_X
 ClientSampleId :
 PB113446ZHE#13


Tot Ion: 75 Resp: 115572
 Ion Ratio Lower Upper
 75 100
 77 1.2 21.4 64.3#



#81
 trans-1,4-Dichloro-2-butene
 Concen: 56.417 ug/l
 RT: 11.14 min Scan# 1638
 Delta R.T. 0.06 min
 Lab File: VX004987.D
 Acq: 03 Oct 2018 23:21

Tgt Ion: 75 Resp: 115572
 Ion Ratio Lower Upper
 75 100
 53 0.2 80.2 120.4#
 89 0.0 37.1 55.7#









#84
 1,2,4-Trimethylbenzene
 Concen: 0.276 ug/l
 RT: 11.81 min Scan# 1748
 Delta R.T. -0.00 min
 Lab File: VX004987.D
 Acq: 03 Oct 2018 23:21

Instrument :
 MSVOA_X
 ClientSampleId :
 PB113446ZHE#13

Tot Ion:105 Resp: 6985
 Ion Ratio Lower Upper
 105 100
 120 45.0 22.8 68.3



#95
 Naphthalene
 Concen: 0.836 ug/l
 RT: 13.83 min Scan# 2080
 Delta R.T. -0.00 min
 Lab File: VX004987.D
 Acq: 03 Oct 2018 23:21

Tgt Ion:128 Resp: 2433
 Ion Ratio Lower Upper
 128 100
 127 13.9 10.4 15.6
 129 13.9 8.7 13.1#

