

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100418\
 Data File : VX004988.D
 Acq On : 03 Oct 2018 23:48
 Operator : JC/MD
 Sample : PB113446ZHE#14
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB113446ZHE#14


Quant Time: Oct 04 01:28:58 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 03 04:00:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	299579	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	464082	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	453922	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	254099	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	215051	50.25	ug/l	0.00
Spiked Amount			Recovery	=	100.50%	
35) Dibromofluoromethane	5.51	113	180035	45.77	ug/l	0.00
Spiked Amount			Recovery	=	91.54%	
50) Toluene-d8	8.72	98	665506	50.90	ug/l	0.00
Spiked Amount			Recovery	=	101.80%	
62) 4-Bromofluorobenzene	11.14	95	245963	52.21	ug/l	0.00
Spiked Amount			Recovery	=	104.42%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
6) Chloroethane	1.73	64	244	Below Cal	#	48
11) Tert butyl alcohol	3.02	59	6294	5.260	ug/l #	73
16) Acetone	2.45	43	22381	8.958	ug/l #	86
18) Methyl Acetate	2.78	43	14238	2.249	ug/l	99
20) Methylene Chloride	2.86	84	5362	0.648	ug/l	89
25) 2-Butanone	4.70	43	4908	1.333	ug/l #	85
29) Tetrahydrofuran	5.18	42	2064	0.902	ug/l	94
31) Cyclohexane	5.66	56	5223	0.959	ug/l #	1
36) 1,1-Dichloropropene	5.67	75	19651	3.662	ug/l #	45
37) Ethyl Acetate	4.70	43	4908	0.689	ug/l #	72
38) Carbon Tetrachloride	5.67	117	24776	4.392	ug/l #	17
43) Isopropyl Acetate	6.46	43	243114	23.646	ug/l	96
48) Methyl methacrylate	7.68	41	8701	1.865	ug/l #	21
51) 4-Methyl-2-Pentanone	8.62	43	4737	0.732	ug/l #	46
52) Toluene	8.79	92	1854	0.201	ug/l	95
54) cis-1,3-Dichloropropene	8.62	75	1630	0.284	ug/l #	58
56) Ethyl methacrylate	9.07	69	122	2.880	ug/l #	1
58) 2-Chloroethyl Vinyl ether	8.34	63	18	14.247	ug/l #	49
59) 2-Hexanone	9.55	43	15728	2.916	ug/l #	53
68) m/p-Xylenes	10.36	106	3868	0.536	ug/l	95
70) Styrene	10.71	104	414	2.172	ug/l #	65
74) N-amyl acetate	10.88	43	87	1.757	ug/l #	40
76) 1,2,3-Trichloropropane	11.14	75	117563	20.642	ug/l #	35
81) trans-1,4-Dichloro-2-buten	11.14	75	117563	56.049	ug/l #	10
84) 1,2,4-Trimethylbenzene	11.80	105	6146	0.209	ug/l	95
85) sec-Butylbenzene	11.80	105	6146	0.383	ug/l #	56
95) Naphthalene	13.84	128	2221	0.823	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.67 min Scan# 740

Delta R.T. 0.01 min

Lab File: VX004988.D

Acq: 03 Oct 2018 23:48

Instrument :

MSVOA_X

ClientSampleId :


PB113446ZHE#14

Tot Ion:168 Resp: 299579

Ion Ratio Lower Upper

168 100

99 49.0 39.3 58.9



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.67

100000

80000

60000



40000

20000

0

Time-->

5.60 5.80 6.00



#6

Chloroethane

Concen: Below Cal

RT: 1.73 min Scan# 94

Delta R.T. 0.01 min

Lab File: VX004988.D


Acq: 03 Oct 2018 23:48

Tgt Ion: 64 Resp: 244

Ion Ratio Lower Upper

64 100

66 58.0 23.8 35.6#



Abundance Ion 63.90 (63.60 to 64.60): VX0

Ion 65.90 (65.60 to 66.60): VX0

1.73


100


50

0

Time-->

1.70 1.75









#11
Tert butyl alcohol
Concen: 5.260 ug/l
RT: 3.02 min Scan# 306
Delta R.T. -0.06 min
Lab File: VX004988.D
Acq: 03 Oct 2018 23:48




Instrument :
MSVOA_X
ClientSampleId :
PB113446ZHE#14


Tot Ion: 59 Resp: 6294
Ion Ratio Lower Upper
59 100
57 0.7 8.6 12.8#



#16
Acetone
Concen: 8.958 ug/l
RT: 2.45 min Scan# 212
Delta R.T. 0.00 min
Lab File: VX004988.D
Acq: 03 Oct 2018 23:48

Tgt Ion: 43 Resp: 22381
Ion Ratio Lower Upper
43 100
58 34.6 21.8 32.6#









#18
Methvl Acetate
Concen: 2.249 ug/l
RT: 2.78 min Scan# 266
Delta R.T. -0.01 min
Lab File: VX004988.D
Acq: 03 Oct 2018 23:48




Instrument :
MSVOA_X
ClientSampleId :
PB113446ZHE#14


Tot Ion: 43 Resp: 14238
Ion Ratio Lower Upper
43 100
74 22.0 17.8 26.8



#20
Methylene Chloride
Concen: 0.648 ug/l
RT: 2.86 min Scan# 280
Delta R.T. 0.01 min
Lab File: VX004988.D
Acq: 03 Oct 2018 23:48

Tgt Ion: 84 Resp: 5362
Ion Ratio Lower Upper
84 100
49 130.5 92.3 138.5
51 40.2 31.8 47.6
86 69.0 48.5 72.7





#25

2-Butanone

Concen: 1.333 ug/l

RT: 4.70 min Scan# 582

Delta R.T. -0.01 min

Lab File: VX004988.D

Acq: 03 Oct 2018 23:48

Instrument :

MSVOA_X

ClientSampleId :


PB113446ZHE#14

Tot Ion: 43 Resp: 4908

Ion Ratio Lower Upper




43 100

72 18.2 20.8 31.2#



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 72.00 (71.70 to 72.70): VX0



#29

Tetrahydrofuran

Concen: 0.902 ug/l

RT: 5.18 min Scan# 660

Delta R.T. 0.02 min

Lab File: VX004988.D

Acq: 03 Oct 2018 23:48


Tgt Ion: 42 Resp: 2064

Ion Ratio Lower Upper

42 100

72 37.6 33.7 50.5



71 42.8 32.2 48.4




Abundance Ion 42.00 (41.70 to 42.70): VX0

Ion 72.00 (71.70 to 72.70): VX0

Ion 71.00 (70.70 to 71.70): VX0





#31

Cyclohexane

Concen: 0.959 ug/l

RT: 5.66 min Scan# 739

Delta R.T. 0.09 min

Lab File: VX004988.D

Acq: 03 Oct 2018 23:48

Instrument :

MSVOA_X

ClientSampleId :

PB113446ZHE#14


Tot Ion: 56 Resp: 5223

Ion Ratio Lower Upper




56 100

69 187.1 25.9 38.9#

84 168.1 71.9 107.9#



Abundance Ion 56.00 (55.70 to 56.70): VX0
Ion 69.00 (68.70 to 69.70): VX0
Ion 84.00 (83.70 to 84.70): VX0



#33

1,2-Dichloroethane-d4

Concen: 50.251 ug/l

RT: 6.07 min Scan# 806

Delta R.T. 0.00 min

Lab File: VX004988.D


Acq: 03 Oct 2018 23:48

Tgt Ion: 65 Resp: 215051



Ion Ratio Lower Upper


65 100

67 52.8 0.0 107.4



Abundance Ion 64.90 (64.60 to 65.60): VX0
Ion 66.90 (66.60 to 67.60): VX0









#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 6.86 min Scan# 936
 Delta R.T. 0.00 min
 Lab File: VX004988.D
 Acq: 03 Oct 2018 23:48

Instrument :
 MSVOA_X
 ClientSampleId :
 PB113446ZHE#14

Tot Ion	Ratio	Lower	Upper
114	100		
63	20.4	0.0	39.8
88	16.0	0.0	27.0




Abundance Ion 113.90 (113.60 to 114.60): V
 Ion 63.00 (62.70 to 63.70): VX0
 Ion 87.90 (87.60 to 88.60): VX0






#35
 Dibromofluoromethane
 Concen: 45.774 ug/l
 RT: 5.51 min Scan# 714
 Delta R.T. 0.01 min
 Lab File: VX004988.D
 Acq: 03 Oct 2018 23:48

Tgt Ion	Ratio	Lower	Upper
113	100		
111	103.5	77.0	115.4
192	23.7	17.0	25.6



Abundance Ion 112.80 (112.50 to 113.50): V
 Ion 110.80 (110.50 to 111.50): V
 Ion 191.70 (191.40 to 192.40): V





#36

1,1-Dichloropropene

Concen: 3.662 ug/l

RT: 5.67 min Scan# 740

Delta R.T. -0.13 min

Lab File: VX004988.D

Acq: 03 Oct 2018 23:48

Instrument :

MSVOA_X

ClientSampleId :

PB113446ZHE#14


Tot Ion: 75 Resp: 19651

Ion Ratio Lower Upper

75 100

110 8.3 20.0 59.9#




77 0.4 27.1 40.7#



Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 109.90 (109.60 to 110.60): VX0

Ion 76.90 (76.60 to 77.60): VX0



#37

Ethyl Acetate

Concen: 0.689 ug/l

RT: 4.70 min Scan# 582

Delta R.T. -0.16 min

Lab File: VX004988.D

Acq: 03 Oct 2018 23:48


Tgt Ion: 43 Resp: 4908

Ion Ratio Lower Upper

43 100

61 2.5 10.4 15.6#



70 0.0 8.7 13.1#




Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 60.90 (60.60 to 61.60): VX0

Ion 70.00 (69.70 to 70.70): VX0





#38

Carbon Tetrachloride

Concen: 4.392 ug/l

RT: 5.67 min Scan# 740

Delta R.T. -0.12 min

Lab File: VX004988.D

Acq: 03 Oct 2018 23:48

Instrument :

MSVOA_X

ClientSampleId :

PB113446ZHE#14


Tot Ion:117 Resp: 24776

Ion Ratio Lower Upper

117 100

119 4.9 75.2 112.8#




121 0.0 22.5 33.7#



Abundance Ion 116.80 (116.50 to 117.50): V

Ion 118.80 (118.50 to 119.50): V

Ion 120.80 (120.50 to 121.50): V



#43

Isopropyl Acetate

Concen: 23.646 ug/l

RT: 6.46 min Scan# 871

Delta R.T. 0.00 min

Lab File: VX004988.D

Acq: 03 Oct 2018 23:48


Tgt Ion: 43 Resp: 243114

Ion Ratio Lower Upper

43 100

61 20.1 14.2 21.4



87 12.6 9.3 13.9




Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 61.00 (60.70 to 61.70): VX0

Ion 87.00 (86.70 to 87.70): VX0









#48
Methyl methacrylate
Concen: 1.865 ug/l
RT: 7.68 min Scan# 1071
Delta R.T. -0.10 min
Lab File: VX004988.D
Acq: 03 Oct 2018 23:48




Instrument :
MSVOA_X
ClientSampleId :
PB113446ZHE#14


Tot Ion: 41 Resp: 8701
Ion Ratio Lower Upper
41 100
69 0.0 63.8 95.6#
39 9.2 40.7 61.1#



#50
Toluene-d8
Concen: 50.896 ug/l
RT: 8.72 min Scan# 1241
Delta R.T. 0.00 min
Lab File: VX004988.D
Acq: 03 Oct 2018 23:48

Tgt Ion: 98 Resp: 665506
Ion Ratio Lower Upper
98 100
100 63.7 51.9 77.9









#51
4-Methyl-2-Pentanone
Concen: 0.732 ug/l
RT: 8.62 min Scan# 1225
Delta R.T. -0.03 min
Lab File: VX004988.D
Acq: 03 Oct 2018 23:48




Instrument :
MSVOA_X
ClientSampleId :
PB113446ZHE#14


Tot Ion: 43 Resp: 4737
Ion Ratio Lower Upper
43 100
58 5.6 30.5 45.7#



#52
Toluene
Concen: 0.201 ug/l
RT: 8.79 min Scan# 1252
Delta R.T. 0.00 min
Lab File: VX004988.D
Acq: 03 Oct 2018 23:48

Tgt Ion: 92 Resp: 1854
Ion Ratio Lower Upper
92 100
91 178.1 136.7 205.1









#54
 cis-1,3-Dichloropropene
 Concen: 0.284 ug/l
 RT: 8.62 min Scan# 1224
 Delta R.T. 0.18 min
 Lab File: VX004988.D
 Acq: 03 Oct 2018 23:48




Instrument :
 MSVOA_X
 ClientSampleId :
 PB113446ZHE#14


Tot Ion: 75 Resp: 1630
 Ion Ratio Lower Upper
 75 100
 77 8.4 25.7 38.5#
 39 20.3 38.3 57.5#



#56
 Ethyl methacrylate
 Concen: 2.880 ug/l
 RT: 9.07 min Scan# 1299
 Delta R.T. -0.11 min
 Lab File: VX004988.D
 Acq: 03 Oct 2018 23:48

Tgt Ion: 69 Resp: 122
 Ion Ratio Lower Upper
 69 100
 41 115.6 56.9 85.3#
 39 177.9 28.4 42.6#









#58
2-Chloroethyl Vinyl ether
Concen: 14.247 ug/l
RT: 8.34 min Scan# 1179
Delta R.T. 0.02 min
Lab File: VX004988.D
Acq: 03 Oct 2018 23:48




Instrument :
MSVOA_X
ClientSampleId :
PB113446ZHE#14


Tot Ion: 63 Resp: 18
Ion Ratio Lower Upper
63 100
106 0.0 21.0 31.6#



#59
2-Hexanone
Concen: 2.916 ug/l
RT: 9.55 min Scan# 1377
Delta R.T. 0.05 min
Lab File: VX004988.D
Acq: 03 Oct 2018 23:48

Tgt Ion: 43 Resp: 15728
Ion Ratio Lower Upper
43 100
58 22.5 28.4 85.3#









#62
4-Bromofluorobenzene
Concen: 52.215 ug/l
RT: 11.14 min Scan# 1638
Delta R.T. 0.00 min
Lab File: VX004988.D
Acq: 03 Oct 2018 23:48




Instrument :
MSVOA_X
ClientSampleId :
PB113446ZHE#14


Tot Ion: 95 Resp: 245963
Ion Ratio Lower Upper
95 100
174 92.5 0.0 185.0
176 89.5 0.0 180.2



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.12 min Scan# 1470
Delta R.T. 0.00 min
Lab File: VX004988.D
Acq: 03 Oct 2018 23:48

Tgt Ion: 117 Resp: 453922
Ion Ratio Lower Upper
117 100
82 54.0 42.2 63.4
119 32.1 27.4 41.0









#68
m/p-Xylenes
Concen: 0.536 ug/l
RT: 10.36 min Scan# 1510
Delta R.T. -0.01 min
Lab File: VX004988.D
Acq: 03 Oct 2018 23:48




Instrument :
MSVOA_X
ClientSampleId :
PB113446ZHE#14


Tot Ion:106 Resp: 3868
Ion Ratio Lower Upper
106 100
91 204.9 157.5 236.3



#70
Styrene
Concen: 2.172 ug/l
RT: 10.71 min Scan# 1568
Delta R.T. 0.00 min
Lab File: VX004988.D
Acq: 03 Oct 2018 23:48

Tgt Ion:104 Resp: 414
Ion Ratio Lower Upper
104 100
78 91.3 40.0 60.0#
103 64.3 44.3 66.5





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.08 min Scan# 1792

Delta R.T. 0.00 min

Lab File: VX004988.D

Acq: 03 Oct 2018 23:48

Instrument :

MSVOA_X

ClientSampleId :

PB113446ZHE#14


Tot Ion:152 Resp: 254099

Ion Ratio Lower Upper

152 100

115 55.6 40.2 120.6




150 157.6 0.0 351.0



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V



#74

N-ethyl acetate

Concen: 1.757 ug/l

RT: 10.88 min Scan# 1596

Delta R.T. -0.02 min

Lab File: VX004988.D

Acq: 03 Oct 2018 23:48

Tgt Ion: 43 Resp: 87


Ion Ratio Lower Upper

43 100

70 0.0 32.6 48.8#

55 59.8 21.6 32.4#

61 0.0 19.1 28.7#






Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 70.00 (69.70 to 70.70): VX0

Ion 55.00 (54.70 to 55.70): VX0

Ion 61.00 (60.70 to 61.70): VX0









#76
 1,2,3-Trichloropropane
 Concen: 20.642 ug/l
 RT: 11.14 min Scan# 1638
 Delta R.T. -0.16 min
 Lab File: VX004988.D
 Acq: 03 Oct 2018 23:48




Instrument :
 MSVOA_X
 ClientSampleId :
 PB113446ZHE#14


Tot Ion: 75 Resp: 117563
 Ion Ratio Lower Upper
 75 100
 77 1.4 21.4 64.3#



#81
 trans-1,4-Dichloro-2-butene
 Concen: 56.049 ug/l
 RT: 11.14 min Scan# 1638
 Delta R.T. 0.06 min
 Lab File: VX004988.D
 Acq: 03 Oct 2018 23:48

Tgt Ion: 75 Resp: 117563
 Ion Ratio Lower Upper
 75 100
 53 0.3 80.2 120.4#
 89 0.0 37.1 55.7#









#84
 1,2,4-Trimethylbenzene
 Concen: 0.209 ug/l
 RT: 11.80 min Scan# 1747
 Delta R.T. -0.01 min
 Lab File: VX004988.D
 Acq: 03 Oct 2018 23:48




Instrument :
 MSVOA_X
 ClientSampleId :
 PB113446ZHE#14




Tot Ion:105 Resp: 6146
 Ion Ratio Lower Upper
 105 100
 120 42.2 22.8 68.3



#85
 sec-Butylbenzene
 Concen: 0.383 ug/l
 RT: 11.80 min Scan# 1747
 Delta R.T. -0.15 min
 Lab File: VX004988.D
 Acq: 03 Oct 2018 23:48

Tgt Ion:105 Resp: 6146
 Ion Ratio Lower Upper
 105 100
 134 0.0 10.2 30.4#





#95
Naphthalene
Concen: 0.823 ug/l
RT: 13.84 min Scan# 2081
Delta R.T. 0.01 min
Lab File: VX004988.D
Acq: 03 Oct 2018 23:48

Instrument :
MSVOA_X
ClientSampleId :
PB113446ZHE#14

Tot Ion	Ratio	Lower	Upper
128	100		
127	13.5	10.4	15.6
129	12.3	8.7	13.1

