

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100418\
 Data File : VX004977.D
 Acq On : 03 Oct 2018 18:55
 Operator : JC/MD
 Sample : PB113446ZHE#03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB113446ZHE#03

Quant Time: Oct 04 02:25:42 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 04 02:08:49 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	3.74	168	269843	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.56	114	425329	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.94	117	412114	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.04	152	234779	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.28	65	182674	47.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.78%	
35) Dibromofluoromethane	3.54	113	162013	44.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.88%	
50) Toluene-d8	8.30	98	609122	50.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.66%	
62) 4-Bromofluorobenzene	11.06	95	227567	52.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.42%	
Target Compounds						
37) Ethyl Acetate	4.93	43	248824	38.114	ug/l #	77
52) Toluene	8.38	92	11539	1.364	ug/l	93
67) Ethyl Benzene	10.22	91	60187	3.587	ug/l #	74
68) m/p-Xylenes	10.22	106	30030	4.580	ug/l	98
69) o-Xylene	10.59	106	10328	1.715	ug/l	98
80) 1,3,5-Trimethylbenzene	11.45	105	17368	1.087	ug/l	96
84) 1,2,4-Trimethylbenzene	11.76	105	55873	3.812	ug/l	99
95) Naphthalene	13.83	128	23127	1.887	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100418\
Data File : VX004977.D
Acq On : 03 Oct 2018 18:55
Operator : JC/MD
Sample : PB113446ZHE#03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB113446ZHE#03

Quant Time: Oct 04 02:25:42 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 04 02:08:49 2018
Response via : Initial Calibration

