

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100418\
 Data File : VX004984.D
 Acq On : 03 Oct 2018 22:01
 Operator : JC/MD
 Sample : PB113446ZHE#10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB113446ZHE#10

Quant Time: Oct 04 03:24:02 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 03 04:00:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	283583	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	440079	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	429300	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	241602	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	194155	47.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.86%	
35) Dibromofluoromethane	5.51	113	166009	44.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.02%	
50) Toluene-d8	8.72	98	629151	50.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.48%	
62) 4-Bromofluorobenzene	11.14	95	231745	51.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.76%	
Target Compounds						
16) Acetone	2.45	43	21245	8.983	ug/l	94
18) Methyl Acetate	2.78	43	13355	2.229	ug/l	99
25) 2-Butanone	4.71	43	5582	1.601	ug/l #	83
43) Isopropyl Acetate	6.46	43	250734	25.718	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100418\
Data File : VX004984.D
Acq On : 03 Oct 2018 22:01
Operator : JC/MD
Sample : PB113446ZHE#10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB113446ZHE#10

Quant Time: Oct 04 03:24:02 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 03 04:00:03 2018
Response via : Initial Calibration

