

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100418\
 Data File : VX004985.D
 Acq On : 03 Oct 2018 22:28
 Operator : JC/MD
 Sample : PB113446ZHE#11
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB113446ZHE#11

Quant Time: Oct 04 03:26:06 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 03 04:00:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	274887	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	463081	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	438397	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	246048	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	198139	50.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.92%	
35) Dibromofluoromethane	5.51	113	175681	44.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.52%	
50) Toluene-d8	8.72	98	642306	49.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.46%	
62) 4-Bromofluorobenzene	11.14	95	237422	50.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.02%	
Target Compounds						
16) Acetone	2.45	43	22678	9.892	ug/l	93
18) Methyl Acetate	2.78	43	14001	2.410	ug/l	96
25) 2-Butanone	4.71	43	5943	1.759	ug/l	98
43) Isopropyl Acetate	6.46	43	250431	24.411	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100418\
Data File : VX004985.D
Acq On : 03 Oct 2018 22:28
Operator : JC/MD
Sample : PB113446ZHE#11
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB113446ZHE#11

Quant Time: Oct 04 03:26:06 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 03 04:00:03 2018
Response via : Initial Calibration

