

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100418\
 Data File : VX004986.D
 Acq On : 03 Oct 2018 22:54
 Operator : JC/MD
 Sample : PB113446ZHE#12
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB113446ZHE#12

Quant Time: Oct 04 03:29:19 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 03 04:00:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	270526	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	437615	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	415187	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	217370	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	193900	50.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.36%	
35) Dibromofluoromethane	5.50	113	169851	45.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.60%	
50) Toluene-d8	8.71	98	615673	49.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.86%	
62) 4-Bromofluorobenzene	11.14	95	221156	49.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.58%	
Target Compounds						
16) Acetone	2.45	43	24819	11.001	ug/l	93
18) Methyl Acetate	2.78	43	11974	2.095	ug/l	96
25) 2-Butanone	4.71	43	4793	1.441	ug/l	100
43) Isopropyl Acetate	6.46	43	249204	25.705	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100418\
Data File : VX004986.D
Acq On : 03 Oct 2018 22:54
Operator : JC/MD
Sample : PB113446ZHE#12
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB113446ZHE#12

Quant Time: Oct 04 03:29:19 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 03 04:00:03 2018
Response via : Initial Calibration

