

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100418\
 Data File : VX005001.D
 Acq On : 04 Oct 2018 06:28
 Operator : JC/MD
 Sample : J5135-14DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SA-TW509-3337DL

Quant Time: Oct 09 07:31:49 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 03 04:00:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	286056	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	422059	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	410377	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	233169	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	199505	48.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.64%	
35) Dibromofluoromethane	5.50	113	177751	49.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.38%	
50) Toluene-d8	8.71	98	605549	50.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.84%	
62) 4-Bromofluorobenzene	11.14	95	223968	52.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.56%	
Target Compounds						
6) Chloroethane	1.72	64	19871	6.176	ug/l	92
12) 1,1-Dichloroethene	2.37	96	5573	1.798	ug/l	94
16) Acetone	2.44	43	1902	0.797	ug/l #	89
24) 1,1-Dichloroethane	3.69	63	179667	26.448	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100418\
Data File : VX005001.D
Acq On : 04 Oct 2018 06:28
Operator : JC/MD
Sample : J5135-14DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SA-TW509-3337DL

Quant Time: Oct 09 07:31:49 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 03 04:00:03 2018
Response via : Initial Calibration

