

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100418\
 Data File : VX005004.D
 Acq On : 04 Oct 2018 07:48
 Operator : JC/MD
 Sample : J5211-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CB-1018-S-TCLP-GRID41-2

Quant Time: Oct 04 09:36:13 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 03 04:00:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	266272	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	428606	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	406605	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	226609	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	198766	52.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.52%	
35) Dibromofluoromethane	5.51	113	175839	48.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.82%	
50) Toluene-d8	8.71	98	587104	48.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.24%	
62) 4-Bromofluorobenzene	11.14	95	217246	49.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.88%	
Target Compounds						
16) Acetone	2.45	43	30265	13.629	ug/l	90
18) Methyl Acetate	2.78	43	7063	1.255	ug/l	98
20) Methylene Chloride	2.86	84	300941	93.944	ug/l	98
43) Isopropyl Acetate	6.46	43	173968	18.321	ug/l	96
95) Naphthalene	13.83	128	5438	1.004	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100418\
Data File : VX005004.D
Acq On : 04 Oct 2018 07:48
Operator : JC/MD
Sample : J5211-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CB-1018-S-TCLP-GRID41-2

Quant Time: Oct 04 09:36:13 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 03 04:00:03 2018
Response via : Initial Calibration

