

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100418\
 Data File : VX005007.D
 Acq On : 04 Oct 2018 09:08
 Operator : JC/MD
 Sample : J5198-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CL-01-100218-A

Quant Time: Oct 04 11:50:18 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 03 04:00:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	275798	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	432514	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	415659	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	232765	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	198624	50.41	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	100.82%	
35) Dibromofluoromethane	5.50	113	172973	47.19	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	94.38%	
50) Toluene-d8	8.72	98	604838	49.63	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	99.26%	
62) 4-Bromofluorobenzene	11.14	95	222955	50.78	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	101.56%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.44	43	83709	36.393	ug/l	94
18) Methyl Acetate	2.78	43	8881	1.524	ug/l	96
20) Methylene Chloride	2.85	84	11303168	3436.703	ug/l	87
30) Chloroform	5.21	83	2444	0.395	ug/l	99
43) Isopropyl Acetate	6.46	43	203956	21.286	ug/l	96
95) Naphthalene	13.83	128	70641	4.316	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100418\
Data File : VX005007.D
Acq On : 04 Oct 2018 09:08
Operator : JC/MD
Sample : J5198-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CL-01-100218-A

Quant Time: Oct 04 11:50:18 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 03 04:00:03 2018
Response via : Initial Calibration

