

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100418\
 Data File : VX005012.D
 Acq On : 04 Oct 2018 11:21
 Operator : JC/MD
 Sample : PB113446ZHE#04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB113446ZHE#04

Quant Time: Oct 04 12:07:05 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 03 04:00:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	252871	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	411568	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	396823	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	217732	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	189148	52.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.72%	
35) Dibromofluoromethane	5.50	113	161087	46.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.36%	
50) Toluene-d8	8.71	98	583550	50.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.64%	
62) 4-Bromofluorobenzene	11.14	95	211302	50.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.16%	
Target Compounds						
16) Acetone	2.45	43	30226	14.333	ug/l	90
18) Methyl Acetate	2.78	43	11973	2.241	ug/l	97
20) Methylene Chloride	2.86	84	6083	1.164	ug/l	90
43) Isopropyl Acetate	6.46	43	226670	24.860	ug/l	98
95) Naphthalene	13.83	128	217065	12.532	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100418\
Data File : VX005012.D
Acq On : 04 Oct 2018 11:21
Operator : JC/MD
Sample : PB113446ZHE#04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB113446ZHE#04

Quant Time: Oct 04 12:07:05 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 03 04:00:03 2018
Response via : Initial Calibration

