

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100418\
 Data File : VX005013.D
 Acq On : 04 Oct 2018 11:48
 Operator : JC/MD
 Sample : PB113446ZHE#05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB113446ZHE#05

Quant Time: Oct 04 12:09:56 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 03 04:00:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	248291	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	408123	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	400204	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	216072	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	181285	51.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.22%	
35) Dibromofluoromethane	5.51	113	154635	44.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.42%	
50) Toluene-d8	8.72	98	583266	50.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.44%	
62) 4-Bromofluorobenzene	11.14	95	211826	51.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.26%	
Target Compounds						
16) Acetone	2.45	43	28700	13.860	ug/l	96
18) Methyl Acetate	2.78	43	11354	2.164	ug/l	100
20) Methylene Chloride	2.85	84	5929	1.149	ug/l	88
25) 2-Butanone	4.71	43	5091	1.668	ug/l	99
43) Isopropyl Acetate	6.46	43	221325	24.479	ug/l	96
95) Naphthalene	13.83	128	102033	6.315	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100418\
Data File : VX005013.D
Acq On : 04 Oct 2018 11:48
Operator : JC/MD
Sample : PB113446ZHE#05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB113446ZHE#05

Quant Time: Oct 04 12:09:56 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 03 04:00:03 2018
Response via : Initial Calibration

