

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100418\
 Data File : VX005031.D
 Acq On : 04 Oct 2018 20:56
 Operator : JC/MD
 Sample : J5207-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 67 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SA-TW507-1519

Quant Time: Oct 05 04:43:57 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 03 04:00:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	224980	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	364830	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	352260	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	193755	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	172142	53.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.12%	
35) Dibromofluoromethane	5.50	113	141761	45.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.70%	
50) Toluene-d8	8.72	98	520611	50.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.30%	
62) 4-Bromofluorobenzene	11.14	95	186386	50.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.66%	
Target Compounds						
16) Acetone	2.45	43	9393	5.006	ug/l	98
95) Naphthalene	13.83	128	2268	0.858	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100418\
Data File : VX005031.D
Acq On : 04 Oct 2018 20:56
Operator : JC/MD
Sample : J5207-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 67 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SA-TW507-1519

Quant Time: Oct 05 04:43:57 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 03 04:00:03 2018
Response via : Initial Calibration

