

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100418\
 Data File : VX005033.D
 Acq On : 04 Oct 2018 21:49
 Operator : JC/MD
 Sample : J5207-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 69 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SA-DUP04-20180927

Quant Time: Oct 05 14:20:32 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 03 04:00:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	223801	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	364455	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	356476	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	201516	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	167482	52.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.78%	
35) Dibromofluoromethane	5.51	113	142152	46.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.04%	
50) Toluene-d8	8.72	98	524760	51.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.20%	
62) 4-Bromofluorobenzene	11.14	95	194708	52.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.26%	
Target Compounds						
16) Acetone	2.45	43	4580	2.454	ug/l #	87
24) 1,1-Dichloroethane	3.70	63	29708	5.590	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100418\
Data File : VX005033.D
Acq On : 04 Oct 2018 21:49
Operator : JC/MD
Sample : J5207-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 69 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SA-DUP04-20180927

Quant Time: Oct 05 14:20:32 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 03 04:00:03 2018
Response via : Initial Calibration

