

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100418\
 Data File : VX005037.D
 Acq On : 04 Oct 2018 23:36
 Operator : JC/MD
 Sample : J5207-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 73 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SA-TW506-4751

Quant Time: Oct 05 05:03:08 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 03 04:00:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	212413	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	338638	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	333246	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	185913	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	158505	52.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.48%	
35) Dibromofluoromethane	5.51	113	134318	46.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.60%	
50) Toluene-d8	8.72	98	490069	51.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.72%	
62) 4-Bromofluorobenzene	11.14	95	176687	51.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.80%	
Target Compounds						
16) Acetone	2.45	43	9235	5.213	ug/l #	80
18) Methyl Acetate	2.78	43	5820	1.297	ug/l	95
32) 1,1,1-Trichloroethane	5.50	97	7368	1.892	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100418\
Data File : VX005037.D
Acq On : 04 Oct 2018 23:36
Operator : JC/MD
Sample : J5207-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 73 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
SA-TW506-4751

Quant Time: Oct 05 05:03:08 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 03 04:00:03 2018
Response via : Initial Calibration

