

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100418\
 Data File : VX005051.D
 Acq On : 05 Oct 2018 06:44
 Operator : JC/MD
 Sample : J5207-13
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 89 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SA-TW510-3236

Quant Time: Oct 05 07:28:04 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 03 04:00:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	213283	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	332971	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	328459	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	184763	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	164851	54.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.22%	
35) Dibromofluoromethane	5.50	113	146720	51.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.98%	
50) Toluene-d8	8.72	98	486355	51.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.68%	
62) 4-Bromofluorobenzene	11.14	95	177555	52.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.06%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
6) Chloroethane	1.72	64	95407	53.047	ug/l	96
12) 1,1-Dichloroethene	2.37	96	36007	15.581	ug/l	92
16) Acetone	2.45	43	7456	4.192	ug/l	93
24) 1,1-Dichloroethane	3.69	63	1293283	255.340	ug/l	98
31) Cyclohexane	5.57	56	1037	0.268	ug/l #	48
32) 1,1,1-Trichloroethane	5.50	97	26662	6.820	ug/l	94
40) Benzene	6.14	78	28235	2.342	ug/l #	84
65) Chlorobenzene	10.14	112	6540	0.776	ug/l	100
73) Isopropylbenzene	11.02	105	5443	0.467	ug/l	95
83) tert-Butylbenzene	11.77	119	11201	1.168	ug/l	85
85) sec-Butylbenzene	11.94	105	50654	4.337	ug/l	83
87) 1,3-Dichlorobenzene	12.03	146	7021	1.112	ug/l	98
88) 1,4-Dichlorobenzene	12.10	146	30610	4.695	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100418\
Data File : VX005051.D
Acq On : 05 Oct 2018 06:44
Operator : JC/MD
Sample : J5207-13
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 89 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SA-TW510-3236

Quant Time: Oct 05 07:28:04 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 03 04:00:03 2018
Response via : Initial Calibration

