

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100418\
 Data File : VX005053.D
 Acq On : 05 Oct 2018 07:37
 Operator : JC/MD
 Sample : J5207-15
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 91 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB04-20180927

Quant Time: Oct 05 09:03:40 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 03 04:00:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	230752	50.00	ug/l	0.01
34) 1,4-Difluorobenzene	6.86	114	335134	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	324879	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	170928	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	164097	49.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.56%	
35) Dibromofluoromethane	5.51	113	153647	54.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.20%	
50) Toluene-d8	8.72	98	484211	51.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.56%	
62) 4-Bromofluorobenzene	11.14	95	166735	49.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.02%	
Target Compounds						
16) Acetone	2.45	43	12091	6.283	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100418\
Data File : VX005053.D
Acq On : 05 Oct 2018 07:37
Operator : JC/MD
Sample : J5207-15
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 91 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleID :
EB04-20180927

Quant Time: Oct 05 09:03:40 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 03 04:00:03 2018
Response via : Initial Calibration

