

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100418\
 Data File : VX005056.D
 Acq On : 05 Oct 2018 08:57
 Operator : JC/MD
 Sample : J5237-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 94 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BSS-UST-1A

Quant Time: Oct 05 11:49:16 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 03 04:00:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	211721	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	334020	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	323519	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	173656	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	168411	55.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.36%	
35) Dibromofluoromethane	5.51	113	143679	50.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.50%	
50) Toluene-d8	8.72	98	473624	50.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.66%	
62) 4-Bromofluorobenzene	11.14	95	168877	49.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.62%	
Target Compounds						
					Qvalue	
16) Acetone	2.45	43	11649	6.597	ug/l #	85
20) Methylene Chloride	2.86	84	3460	0.517	ug/l	99
25) 2-Butanone	4.70	43	2794	1.073	ug/l #	89
29) Tetrahydrofuran	5.17	42	1627	1.006	ug/l #	88
30) Chloroform	5.22	83	11069	2.329	ug/l #	80
47) Bromodichloromethane	7.90	83	2309	0.662	ug/l	98
60) Dibromochloromethane	9.59	129	663	0.228	ug/l	93
95) Naphthalene	13.83	128	2056	0.860	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100418\
Data File : VX005056.D
Acq On : 05 Oct 2018 08:57
Operator : JC/MD
Sample : J5237-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 94 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BSS-UST-1A

Quant Time: Oct 05 11:49:16 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 03 04:00:03 2018
Response via : Initial Calibration

