

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\WX100418\
 Data File : VX005016.D
 Acq On : 04 Oct 2018 13:08
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 04 13:52:23 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 03 04:00:03 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	83	0.00
2 T	Dichlorodifluoromethane	0.682	0.694	-1.8	71	0.00
3 P	Chloromethane	0.827	0.776	6.2	72	0.00
4 C	Vinyl Chloride	0.776	0.781	-0.6#	78	0.00
5 T	Bromomethane	0.460	0.527	-14.6	91	0.00
6 T	Chloroethane	0.495	0.457	7.7	83	0.00
7 T	Trichlorofluoromethane	0.960	1.003	-4.5	78	0.00
8 T	Diethyl Ether	0.417	0.432	-3.6	88	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.547	0.578	-5.7	90	0.00
10 T	Methyl Iodide	0.732	0.689	5.9	66	0.00
11 T	Tert butyl alcohol	0.200	0.201	-0.5	82	0.00
12 CM	1,1-Dichloroethene	0.542	0.557	-2.8#	88	0.00
13 T	Acrolein	0.142	0.078	45.1#	46#	0.00
14 T	Allyl chloride	1.182	1.238	-4.7	84	0.00
15 T	Acrylonitrile	0.440	0.463	-5.2	83	0.00
16 T	Acetone	0.417	0.395	5.3	75	0.00
17 T	Carbon Disulfide	1.628	1.506	7.5	74	0.00
18 T	Methyl Acetate	1.056	1.291	-22.3#	101	0.00
19 T	Methyl tert-butyl Ether	1.869	2.130	-14.0	95	0.00
20 T	Methylene Chloride	0.670	0.660	1.5	79	0.00
21 T	trans-1,2-Dichloroethene	0.593	0.600	-1.2	80	0.00
22 T	Diisopropyl ether	1.991	2.166	-8.8	87	0.00
23 T	Vinyl Acetate	1.822	1.871	-2.7	79	0.00
24 P	1,1-Dichloroethane	1.187	1.171	1.3	78	0.00
25 T	2-Butanone	0.615	0.619	-0.7	81	0.00
26 T	2,2-Dichloropropane	0.827	0.858	-3.7	86	0.00
27 T	cis-1,2-Dichloroethene	0.663	0.677	-2.1	81	0.00
28 T	Bromochloromethane	0.559	0.536	4.1	81	0.00
29 T	Tetrahydrofuran	0.382	0.406	-6.3	82	0.00
30 C	Chloroform	1.122	1.116	0.5#	85	0.00
31 T	Cyclohexane	0.909	1.004	-10.5	82	0.00
32 T	1,1,1-Trichloroethane	0.916	0.943	-2.9	84	0.00
33 S	1,2-Dichloroethane-d4	0.714	0.692	3.1	75	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	87	0.00
35 S	Dibromofluoromethane	0.424	0.373	12.0	73	0.00
36 T	1,1-Dichloropropene	0.578	0.538	6.9	83	0.00
37 T	Ethyl Acetate	0.767	0.715	6.8	81	0.00
38 T	Carbon Tetrachloride	0.608	0.530	12.8	78	0.00
39 T	Methylcyclohexane	0.590	0.630	-6.8	95	0.00
40 TM	Benzene	1.810	1.713	5.4	86	0.00
41 T	Methacrylonitrile	0.426	0.407	4.5	80	0.00
42 TM	1,2-Dichloroethane	0.640	0.586	8.4	82	0.00
43 T	Isopropyl Acetate	1.108	1.105	0.3	81	0.00
44 TM	Trichloroethene	0.445	0.449	-0.9	97	0.00
45 C	1,2-Dichloropropane	0.444	0.449	-1.1#	95	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\WX100418\
 Data File : VX005016.D
 Acq On : 04 Oct 2018 13:08
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 04 13:52:23 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 03 04:00:03 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.279	0.283	-1.4	96	0.00
47 T	Bromodichloromethane	0.522	0.544	-4.2	96	0.00
48 T	Methyl methacrylate	0.503	0.566	-12.5	97	0.00
49 T	1,4-Dioxane	0.012	0.012	0.0	81	0.00
50 S	Toluene-d8	1.409	1.425	-1.1	94	0.00
51 T	4-Methyl-2-Pentanone	0.697	0.768	-10.2	96	0.00
52 CM	Toluene	0.995	1.035	-4.0#	91	0.00
53 T	t-1,3-Dichloropropene	0.590	0.635	-7.6	89	0.00
54 T	cis-1,3-Dichloropropene	0.618	0.680	-10.0	97	0.00
55 T	1,1,2-Trichloroethane	0.436	0.433	0.7	87	0.00
56 T	Ethyl methacrylate	0.628	0.710	-13.1	94	0.00
57 T	1,3-Dichloropropane	0.725	0.733	-1.1	86	0.00
58 T	2-Chloroethyl Vinyl ether	0.331	0.384	-16.0	95	0.00
59 T	2-Hexanone	0.581	0.608	-4.6	85	0.00
60 T	Dibromochloromethane	0.436	0.444	-1.8	82	0.00
61 T	1,2-Dibromoethane	0.467	0.451	3.4	80	0.00
62 S	4-Bromofluorobenzene	0.508	0.532	-4.7	98	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	92	0.00
64 T	Tetrachloroethene	0.492	0.493	-0.2	91	0.00
65 PM	Chlorobenzene	1.282	1.247	2.7	91	0.00
66 T	1,1,1,2-Tetrachloroethane	0.451	0.450	0.2	94	0.00
67 C	Ethyl Benzene	2.036	2.115	-3.9#	97	0.00
68 T	m/p-Xylenes	0.795	0.828	-4.2	93	0.00
69 T	o-Xylene	0.731	0.808	-10.5	100	0.00
70 T	Styrene	1.227	1.357	-10.6	98	0.00
71 P	Bromoform	0.386	0.389	-0.8	91	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
73 T	Isopropylbenzene	3.152	3.540	-12.3	99	0.00
74 T	N-amyl acetate	1.531	1.761	-15.0	101	0.00
75 P	1,1,2,2-Tetrachloroethane	1.283	1.247	2.8	97	0.00
76 T	1,2,3-Trichloropropane	1.121	1.179	-5.2	100	0.00
77 T	Bromobenzene	0.915	0.951	-3.9	98	0.00
78 T	n-propylbenzene	3.616	4.028	-11.4	98	0.00
79 T	2-Chlorotoluene	2.254	2.422	-7.5	98	0.00
80 T	1,3,5-Trimethylbenzene	2.664	3.007	-12.9	98	0.00
81 T	trans-1,4-Dichloro-2-butene	0.367	0.378	-3.0	97	0.00
82 T	4-Chlorotoluene	2.653	2.863	-7.9	97	0.00
83 T	tert-Butylbenzene	2.595	2.939	-13.3	101	0.00
84 T	1,2,4-Trimethylbenzene	2.728	3.181	-16.6	101	0.00
85 T	sec-Butylbenzene	3.161	3.496	-10.6	98	0.00
86 T	p-Isopropyltoluene	2.863	3.178	-11.0	98	0.00
87 T	1,3-Dichlorobenzene	1.709	1.740	-1.8	97	0.00
88 T	1,4-Dichlorobenzene	1.765	1.776	-0.6	98	0.00
89 T	n-Butylbenzene	2.601	2.838	-9.1	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX100418\
Data File : VX005016.D
Acq On : 04 Oct 2018 13:08
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Oct 04 13:52:23 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 03 04:00:03 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.477	0.482	-1.0	100	0.00
91 T	1,2-Dichlorobenzene	1.786	1.767	1.1	96	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.282	0.301	-6.7	104	0.00
93 T	1,2,4-Trichlorobenzene	1.253	1.328	-6.0	99	0.00
94 T	Hexachlorobutadiene	0.650	0.626	3.7	97	0.00
95 T	Naphthalene	3.727	4.438	-19.1	103	0.00
96 T	1,2,3-Trichlorobenzene	1.294	1.356	-4.8	98	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6