

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100419\
 Data File : VX012809.D
 Acq On : 04 Oct 2019 14:15
 Operator : JC/SP
 Sample : K5189-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SA-PZ163I-20190930

Quant Time: Oct 05 05:24:55 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 23 12:27:01 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	178037	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	295894	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	251320	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	92948	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	115832	44.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.36%	
35) Dibromofluoromethane	5.49	113	89203	50.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.24%	
50) Toluene-d8	8.71	98	350811	51.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.38%	
62) 4-Bromofluorobenzene	11.14	95	121919	45.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.38%	
Target Compounds						
16) Acetone	2.44	43	6215	4.526	ug/l	96
64) Tetrachloroethene	9.33	164	3529	2.014	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100419\
Data File : VX012809.D
Acq On : 04 Oct 2019 14:15
Operator : JC/SP
Sample : K5189-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SA-PZ163I-20190930

Quant Time: Oct 05 05:24:55 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 23 12:27:01 2019
Response via : Initial Calibration

