

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100419\
 Data File : VX012812.D
 Acq On : 04 Oct 2019 15:25
 Operator : JC/SP
 Sample : K5189-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SA-MW127D-20190930

Quant Time: Oct 05 05:31:26 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 23 12:27:01 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	176099	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	302004	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	256439	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	98662	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	116722	45.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.04%	
35) Dibromofluoromethane	5.49	113	91929	51.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.22%	
50) Toluene-d8	8.71	98	355454	51.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.64%	
62) 4-Bromofluorobenzene	11.14	95	123164	45.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.44%	
Target Compounds						
3) Chloromethane	1.32	50	1237	0.544	ug/l #	87
16) Acetone	2.43	43	6184	4.553	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100419\
Data File : VX012812.D
Acq On : 04 Oct 2019 15:25
Operator : JC/SP
Sample : K5189-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SA-MW127D-20190930

Quant Time: Oct 05 05:31:26 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 23 12:27:01 2019
Response via : Initial Calibration

