

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX100419\
 Data File : VX012813.D
 Acq On : 04 Oct 2019 15:48
 Operator : JC/SP
 Sample : K5189-17
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SA-PZ138I-20191001

Quant Time: Oct 05 05:33:46 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 23 12:27:01 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	178134	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	310225	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	260330	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	102478	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	121378	46.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.58%	
35) Dibromofluoromethane	5.49	113	92805	50.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.46%	
50) Toluene-d8	8.71	98	361907	50.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.74%	
62) 4-Bromofluorobenzene	11.14	95	124920	44.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.30%	
Target Compounds						
3) Chloromethane	1.32	50	1925	0.837	ug/l #	72
16) Acetone	2.44	43	7705	5.608	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX100419\
Data File : VX012813.D
Acq On : 04 Oct 2019 15:48
Operator : JC/SP
Sample : K5189-17
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SA-PZ138I-20191001

Quant Time: Oct 05 05:33:46 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 23 12:27:01 2019
Response via : Initial Calibration

