

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100421\
 Data File : VX024521.D
 Acq On : 04 Oct 2021 15:34
 Operator : JC/MD
 Sample : M3911-15RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP-BLANKRE

Quant Time: Oct 05 03:20:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	159421	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	247828	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	225672	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	118548	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	112331	50.00	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 100.00%	
35) Dibromofluoromethane	5.397	113	91160	55.37	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 110.74%	
50) Toluene-d8	8.653	98	338013	55.78	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 111.56%	
62) 4-Bromofluorobenzene	11.085	95	127114	53.35	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 106.70%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100421\
Data File : VX024521.D
Acq On : 04 Oct 2021 15:34
Operator : JC/MD
Sample : M3911-15RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANKRE

Quant Time: Oct 05 03:20:37 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

