

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100421\
 Data File : VX024522.D
 Acq On : 04 Oct 2021 15:58
 Operator : JC/MD
 Sample : M4063-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 P287-DITCH

Quant Time: Oct 05 03:20:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	142688	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	234079	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	252358	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	116322	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	102746	51.09	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.18%
35) Dibromofluoromethane	5.397	113	91820	59.05	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	118.10%
50) Toluene-d8	8.653	98	333159	58.21	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	116.42%#
62) 4-Bromofluorobenzene	11.085	95	125532	55.78	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	111.56%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100421\
 Data File : VX024522.D
 Acq On : 04 Oct 2021 15:58
 Operator : JC/MD
 Sample : M4063-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
ClientSampled :
 P287-DITCH

Quant Time: Oct 05 03:20:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

