

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100421\
 Data File : VX024524.D
 Acq On : 04 Oct 2021 16:45
 Operator : JC/MD
 Sample : M4063-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SW-RR-01-(2.0)

Quant Time: Oct 05 03:21:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	138217	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	237867	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	230208	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	100181	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	104596	53.69	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 107.38%	
35) Dibromofluoromethane	5.397	113	82004	51.89	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.78%	
50) Toluene-d8	8.653	98	291165	50.06	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.12%	
62) 4-Bromofluorobenzene	11.085	95	108008	47.23	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 94.46%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100421\
Data File : VX024524.D
Acq On : 04 Oct 2021 16:45
Operator : JC/MD
Sample : M4063-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SW-RR-01-(2.0)

Quant Time: Oct 05 03:21:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

