

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100421\
 Data File : VX024533.D
 Acq On : 04 Oct 2021 20:17
 Operator : JC/MD
 Sample : M3989-05 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-7

Quant Time: Oct 05 03:23:00 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	164072	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	241649	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	230504	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	104759	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	112936	48.84	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 97.68%			
35) Dibromofluoromethane	5.397	113	93282	58.11	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 116.22%			
50) Toluene-d8	8.653	98	295633	50.04	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 100.08%			
62) 4-Bromofluorobenzene	11.085	95	109037	46.94	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 93.88%			
Target Compounds						
						Qvalue
11) Tert butyl alcohol	2.971	59	48241	81.36	ug/l	99
17) Carbon Disulfide	2.514	76	2328	0.83	ug/l	100
19) Methyl tert-butyl Ether	3.123	73	279254	49.27	ug/l	99
22) Diisopropyl ether	3.770	45	4174	0.74	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100421\
Data File : VX024533.D
Acq On : 04 Oct 2021 20:17
Operator : JC/MD
Sample : M3989-05 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-7

Quant Time: Oct 05 03:23:00 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

