

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100421\
 Data File : VX024534.D
 Acq On : 04 Oct 2021 20:40
 Operator : JC/MD
 Sample : M3989-02 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-2R

Quant Time: Oct 05 03:23:11 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	149391	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	264044	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	250817	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	95343	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	106297	50.49	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 100.98%		
35) Dibromofluoromethane	5.397	113	88547	50.48	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 100.96%		
50) Toluene-d8	8.653	98	290321	44.97	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 89.94%#		
62) 4-Bromofluorobenzene	11.085	95	119072	46.91	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 93.82%		
Target Compounds						Qvalue
19) Methyl tert-butyl Ether	3.123	73	2066	0.40	ug/l	# 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100421\
Data File : VX024534.D
Acq On : 04 Oct 2021 20:40
Operator : JC/MD
Sample : M3989-02 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-2R

Quant Time: Oct 05 03:23:11 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

