

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100421\
 Data File : VX024539.D
 Acq On : 04 Oct 2021 23:48
 Operator : JC/MD
 Sample : VX1004WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1004WBL02

Quant Time: Oct 05 03:24:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	139987	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	239344	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	235767	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	105844	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	104038	52.73	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.46%
35) Dibromofluoromethane	5.397	113	80827	50.83	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.66%
50) Toluene-d8	8.653	98	295531	50.50	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.00%
62) 4-Bromofluorobenzene	11.085	95	112400	48.85	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.70%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100421\
Data File : VX024539.D
Acq On : 04 Oct 2021 23:48
Operator : JC/MD
Sample : VX1004WBL02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1004WBL02

Quant Time: Oct 05 03:24:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

