

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100421\
 Data File : VX024542.D
 Acq On : 05 Oct 2021 00:58
 Operator : JC/MD
 Sample : PB139556TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB139556TB

Quant Time: Oct 05 08:33:41 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	150194	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	222329	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	255225	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	90820	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	103461	48.88	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.76%
35) Dibromofluoromethane	5.397	113	85918	58.17	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	116.34%
50) Toluene-d8	8.653	98	329715	60.65	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	121.30%#
62) 4-Bromofluorobenzene	11.085	95	110463	51.68	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.36%
Target Compounds						
					Qvalue	
20) Methylene Chloride	2.794	84	10266	5.46	ug/l	95
25) 2-Butanone	4.574	43	15166	10.40	ug/l	91
43) Isopropyl Acetate	6.348	43	19293	4.77	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100421\
Data File : VX024542.D
Acq On : 05 Oct 2021 00:58
Operator : JC/MD
Sample : PB139556TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB139556TB

Quant Time: Oct 05 08:33:41 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

