

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100421\
 Data File : VX024543.D
 Acq On : 05 Oct 2021 01:21
 Operator : JC/MD
 Sample : PB139574TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB139574TB

Quant Time: Oct 05 08:33:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	157148	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	239713	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	280228	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	125468	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	113420	51.21	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 102.42%			
35) Dibromofluoromethane	5.397	113	92700	58.21	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 116.42%			
50) Toluene-d8	8.653	98	306312	52.26	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 104.52%			
62) 4-Bromofluorobenzene	11.085	95	123622	53.64	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 107.28%			
Target Compounds					Qvalue	
16) Acetone	2.386	43	14244	5.32	ug/l	93
20) Methylene Chloride	2.794	84	8596	4.14	ug/l	92
43) Isopropyl Acetate	6.355	43	20859	4.78	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100421\
Data File : VX024543.D
Acq On : 05 Oct 2021 01:21
Operator : JC/MD
Sample : PB139574TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB139574TB

Quant Time: Oct 05 08:33:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

