

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100421\
 Data File : VX024544.D
 Acq On : 05 Oct 2021 01:45
 Operator : JC/MD
 Sample : PB139556ZHE#01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB139556ZHE#01

Quant Time: Oct 05 08:34:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	140903	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	271595	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	247755	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	116322	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	105603	53.18	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	106.36%
35) Dibromofluoromethane	5.397	113	82674	45.82	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.64%
50) Toluene-d8	8.653	98	343937	51.79	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	103.58%
62) 4-Bromofluorobenzene	11.085	95	118719	45.47	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.94%
Target Compounds						
20) Methylene Chloride	2.794	84	9473	5.35	ug/l	94
25) 2-Butanone	4.568	43	14207	10.39	ug/l #	88
43) Isopropyl Acetate	6.349	43	20953	4.24	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100421\
Data File : VX024544.D
Acq On : 05 Oct 2021 01:45
Operator : JC/MD
Sample : PB139556ZHE#01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB139556ZHE#01

Quant Time: Oct 05 08:34:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

