

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100421\
 Data File : VX024545.D
 Acq On : 05 Oct 2021 02:08
 Operator : JC/MD
 Sample : PB139556ZHE#02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB139556ZHE#02

Quant Time: Oct 05 08:34:18 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	135071	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	258467	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	251147	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	102766	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	106459	55.92	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	111.84%		
35) Dibromofluoromethane	5.397	113	79173	46.11	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	92.22%		
50) Toluene-d8	8.653	98	305658	48.37	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	96.74%		
62) 4-Bromofluorobenzene	11.085	95	110378	44.42	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	88.84%		
Target Compounds						Qvalue
20) Methylene Chloride	2.794	84	8590	5.00	ug/l	91
25) 2-Butanone	4.574	43	12927	9.86	ug/l #	87
43) Isopropyl Acetate	6.348	43	20840	4.43	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100421\
Data File : VX024545.D
Acq On : 05 Oct 2021 02:08
Operator : JC/MD
Sample : PB139556ZHE#02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB139556ZHE#02

Quant Time: Oct 05 08:34:18 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

