

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100421\
 Data File : VX024547.D
 Acq On : 05 Oct 2021 02:55
 Operator : JC/MD
 Sample : PB139556ZHE#04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB139556ZHE#04

Quant Time: Oct 05 08:34:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	132575	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	268477	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	261970	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	91903	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	105001	56.20	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 112.40%			
35) Dibromofluoromethane	5.397	113	89390	50.12	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 100.24%			
50) Toluene-d8	8.653	98	313082	47.69	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 95.38%			
62) 4-Bromofluorobenzene	11.085	95	112691	43.66	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 87.32%			
Target Compounds					Qvalue	
16) Acetone	2.386	43	14733	8.91	ug/l	92
20) Methylene Chloride	2.794	84	6910	3.89	ug/l	96
43) Isopropyl Acetate	6.355	43	21293	4.36	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100421\
Data File : VX024547.D
Acq On : 05 Oct 2021 02:55
Operator : JC/MD
Sample : PB139556ZHE#04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB139556ZHE#04

Quant Time: Oct 05 08:34:38 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

