

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100421\
 Data File : VX024556.D
 Acq On : 05 Oct 2021 06:26
 Operator : JC/MD
 Sample : M3989-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-4

Quant Time: Oct 05 08:36:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	155550	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	270876	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	245685	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	107787	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	112555	51.34	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 102.68%	
35) Dibromofluoromethane	5.397	113	90353	50.21	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.42%	
50) Toluene-d8	8.653	98	286099	43.20	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 86.40%#	
62) 4-Bromofluorobenzene	11.085	95	114891	44.12	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 88.24%	
Target Compounds						
19) Methyl tert-butyl Ether	3.123	73	858446	159.75	ug/l	Qvalue 100
29) Tetrahydrofuran	5.025	42	2592	2.69	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100421\
Data File : VX024556.D
Acq On : 05 Oct 2021 06:26
Operator : JC/MD
Sample : M3989-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-4

Quant Time: Oct 05 08:36:05 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

