

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100421\
 Data File : VX024558.D
 Acq On : 05 Oct 2021 07:13
 Operator : JC/MD
 Sample : M3989-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-8

Quant Time: Oct 05 08:36:23 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	126749	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	218083	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	246421	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	104212	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	96344	53.93	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 107.86%	
35) Dibromofluoromethane	5.397	113	75542	52.14	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.28%	
50) Toluene-d8	8.653	98	268487	50.35	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.70%	
62) 4-Bromofluorobenzene	11.085	95	112491	53.66	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 107.32%	
Target Compounds						Qvalue
16) Acetone	2.386	43	8884	1.72	ug/l	97
19) Methyl tert-butyl Ether	3.123	73	71296	16.28	ug/l	99
67) Ethyl Benzene	10.195	91	3257	0.36	ug/l	91
68) m/p-Xylenes	10.311	106	977	0.28	ug/l	96
70) Styrene	10.665	104	1778	0.31	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100421\
Data File : VX024558.D
Acq On : 05 Oct 2021 07:13
Operator : JC/MD
Sample : M3989-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-8

Quant Time: Oct 05 08:36:23 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

