

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100421\
 Data File : VX024524.D
 Acq On : 04 Oct 2021 16:45
 Operator : JC/MD
 Sample : M4063-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SW-RR-01-(2.0)

Quant Time: Oct 05 03:21:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	138217	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	237867	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	230208	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	100181	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	104596	53.69	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 107.38%	
35) Dibromofluoromethane	5.397	113	82004	51.89	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.78%	
50) Toluene-d8	8.653	98	291165	50.06	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.12%	
62) 4-Bromofluorobenzene	11.085	95	108008	47.23	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 94.46%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100421\
Data File : VX024524.D
Acq On : 04 Oct 2021 16:45
Operator : JC/MD
Sample : M4063-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SW-RR-01-(2.0)

Quant Time: Oct 05 03:21:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.00 ug/l
RT: 5.562 min Scan# 734
Delta R.T. 0.006 min
Lab File: VX024524.D
Acq: 04 Oct 2021 16:45




Instrument :
MSVOA_X
ClientSampleId :
SW-RR-01-(2.0)


Tgt Ion:168 Resp: 138217
Ion Ratio Lower Upper
168 100
99 59.8 45.8 68.6



#33
1,2-Dichloroethane-d4
Concen: 53.69 ug/l
RT: 5.964 min Scan# 800
Delta R.T. 0.000 min
Lab File: VX024524.D
Acq: 04 Oct 2021 16:45

Tgt Ion: 65 Resp: 104596
Ion Ratio Lower Upper
65 100
67 49.6 0.0 105.4









#34
1,4-Difluorobenzene
Concen: 50.00 ug/l
RT: 6.769 min Scan# 932
Delta R.T. -0.000 min
Lab File: VX024524.D
Acq: 04 Oct 2021 16:45




Instrument :
MSVOA_X
ClientSampleId :
SW-RR-01-(2.0)


Tgt Ion:114 Resp: 237867
Ion Ratio Lower Upper
114 100
63 20.3 0.0 40.8
88 16.6 0.0 32.8



#35
Dibromofluoromethane
Concen: 51.89 ug/l
RT: 5.397 min Scan# 707
Delta R.T. 0.000 min
Lab File: VX024524.D
Acq: 04 Oct 2021 16:45

Tgt Ion:113 Resp: 82004
Ion Ratio Lower Upper
113 100
111 103.5 81.9 122.9
192 19.4 16.7 25.1









#50
Toluene-d8
Concen: 50.06 ug/l
RT: 8.653 min Scan# 1241
Delta R.T. -0.000 min
Lab File: VX024524.D
Acq: 04 Oct 2021 16:45




Instrument :
MSVOA_X
ClientSampleId :
SW-RR-01-(2.0)


Tgt Ion: 98 Resp: 291165
Ion Ratio Lower Upper
98 100
100 64.0 51.9 77.9



#62
4-Bromofluorobenzene
Concen: 47.23 ug/l
RT: 11.085 min Scan# 1640
Delta R.T. 0.000 min
Lab File: VX024524.D
Acq: 04 Oct 2021 16:45

Tgt Ion: 95 Resp: 108008
Ion Ratio Lower Upper
95 100
174 80.6 0.0 154.6
176 77.9 0.0 155.8









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX024524.D
Acq: 04 Oct 2021 16:45

Instrument :
MSVOA_X
ClientSampleId :
SW-RR-01-(2.0)

Tgt Ion:117 Resp: 230208
Ion Ratio Lower Upper
117 100
82 58.2 47.4 71.2
119 31.6 26.6 39.8



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX024524.D
Acq: 04 Oct 2021 16:45

Tgt Ion:152 Resp: 100181
Ion Ratio Lower Upper
152 100
115 60.9 41.1 123.3
150 158.7 0.0 349.0

