



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100421\
 Data File : VX024510.D
 Acq On : 04 Oct 2021 10:21
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Title : SW846 8260
 Last Update : Thu Sep 23 18:24:34 2021



AutoFind: Scans 1639, 1640, 1641; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.5	7042	PASS
75	95	30	60	55.3	19965	PASS
95	95	100	100	100.0	36096	PASS
96	95	5	9	7.1	2567	PASS
173	174	0.00	2	1.2	353	PASS
174	95	50	100	82.8	29883	PASS
175	174	5	9	7.2	2139	PASS
176	174	95	101	96.7	28893	PASS
177	176	5	9	6.5	1867	PASS