

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100422\
 Data File : VX031880.D
 Acq On : 04 Oct 2022 20:12
 Operator : JC/MD
 Sample : VSTDC0050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDC0050EC

Quant Time: Oct 05 06:28:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100322W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 04 06:59:30 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	139171	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	191077	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	206110	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	158486	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	92427	37.070	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	74.140%
35) Dibromofluoromethane	5.385	113	99128	43.665	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	87.320%
50) Toluene-d8	8.653	98	348746	42.942	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	85.880%#
62) 4-Bromofluorobenzene	11.079	95	127682	43.659	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.320%

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.167	85	78719	43.753	ug/l	98
3) Chloromethane	1.288	50	62851	33.324	ug/l	97
4) Vinyl Chloride	1.368	62	84413	35.135	ug/l	98
5) Bromomethane	1.612	94	70551	45.148	ug/l	98
6) Chloroethane	1.685	64	79371	38.376	ug/l	97
7) Trichlorofluoromethane	1.886	101	171090	34.895	ug/l	98
8) Diethyl Ether	2.130	74	55294	41.833	ug/l	74
9) 1,1,2-Trichlorotrifluo...	2.325	101	86811	40.986	ug/l #	81
10) Methyl Iodide	2.453	142	103280	51.945	ug/l	96
11) Tert butyl alcohol	3.008	59	72692	209.410	ug/l #	81
12) 1,1-Dichloroethene	2.319	96	81608	41.960	ug/l #	75
13) Acrolein	2.240	56	44226	215.195	ug/l	99
14) Allyl chloride	2.660	41	86570	34.871	ug/l #	91
15) Acrylonitrile	3.069	53	179388	197.840	ug/l	97
16) Acetone	2.392	43	133524	229.472	ug/l #	85
17) Carbon Disulfide	2.508	76	172479	37.750	ug/l	98
18) Methyl Acetate	2.703	43	92909	39.836	ug/l #	82
19) Methyl tert-butyl Ether	3.111	73	258281	40.793	ug/l	94
20) Methylene Chloride	2.788	84	88874	40.745	ug/l #	77
21) trans-1,2-Dichloroethene	3.093	96	92362	40.509	ug/l #	80
22) Diisopropyl ether	3.764	45	198844	37.774	ug/l #	89
23) Vinyl Acetate	3.721	43	778460	197.697	ug/l #	85
24) 1,1-Dichloroethane	3.611	63	136363	38.546	ug/l	95
25) 2-Butanone	4.562	43	219723	192.051	ug/l #	81
26) 2,2-Dichloropropane	4.477	77	101388	33.570	ug/l	89
27) cis-1,2-Dichloroethene	4.489	96	109385	41.261	ug/l	75
28) Bromochloromethane	4.898	49	53346	37.312	ug/l #	34
29) Tetrahydrofuran	5.013	42	140260	198.555	ug/l #	74
30) Chloroform	5.093	83	165371	39.075	ug/l	99
31) Cyclohexane	5.471	56	112799	37.986	ug/l #	79
32) 1,1,1-Trichloroethane	5.385	97	153720	38.276	ug/l #	92
36) 1,1-Dichloropropene	5.696	75	113617	40.362	ug/l	90
37) Ethyl Acetate	4.715	43	87271	40.957	ug/l #	90
38) Carbon Tetrachloride	5.678	117	140756	40.279	ug/l	95
39) Methylcyclohexane	7.379	83	148656	43.047	ug/l #	83
40) Benzene	6.038	78	346375	42.595	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100422\
 Data File : VX031880.D
 Acq On : 04 Oct 2022 20:12
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC

Quant Time: Oct 05 06:28:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100322W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 04 06:59:30 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.922	41	47238	40.797	ug/l #	86
42) 1,2-Dichloroethane	6.086	62	113077	39.692	ug/l	91
43) Isopropyl Acetate	6.342	43	144291	41.881	ug/l #	87
44) Trichloroethene	7.129	130	122007	46.441	ug/l	81
45) 1,2-Dichloropropane	7.434	63	77180	40.512	ug/l	99
46) Dibromomethane	7.580	93	68596	43.198	ug/l #	76
47) Bromodichloromethane	7.824	83	123280	38.791	ug/l	97
48) Methyl methacrylate	7.696	41	66692	43.258	ug/l #	79
49) 1,4-Dioxane	7.696	88	43335	867.218	ug/l #	82
51) 4-Methyl-2-Pentanone	8.574	43	455122	210.448	ug/l	88
52) Toluene	8.720	92	243547	42.720	ug/l	98
53) t-1,3-Dichloropropene	8.976	75	120485	39.758	ug/l	100
54) cis-1,3-Dichloropropene	8.366	75	134983	40.649	ug/l #	84
55) 1,1,2-Trichloroethane	9.153	97	100238	44.113	ug/l	95
56) Ethyl methacrylate	9.116	69	126664	48.209	ug/l #	80
57) 1,3-Dichloropropane	9.311	76	149650	42.407	ug/l	97
58) 2-Chloroethyl Vinyl ether	8.245	63	302427	223.373	ug/l #	79
59) 2-Hexanone	9.433	43	339620	216.015	ug/l	85
60) Dibromochloromethane	9.525	129	114902	41.653	ug/l	98
61) 1,2-Dibromoethane	9.610	107	111566	44.215	ug/l	97
64) Tetrachloroethene	9.275	164	126951	49.268	ug/l	91
65) Chlorobenzene	10.080	112	286485	45.841	ug/l	98
66) 1,1,1,2-Tetrachloroethane	10.165	131	108365	45.049	ug/l	97
67) Ethyl Benzene	10.195	91	465068	45.208	ug/l	95
68) m/p-Xylenes	10.299	106	396107	93.266	ug/l	87
69) o-Xylene	10.640	106	196530	48.032	ug/l	86
70) Styrene	10.659	104	319702	48.072	ug/l	93
71) Bromoform	10.799	173	86905	42.961	ug/l #	100
73) Isopropylbenzene	10.964	105	500006	44.665	ug/l	97
74) N-amyl acetate	10.842	43	122246	46.625	ug/l #	84
75) 1,1,2,2-Tetrachloroethane	11.214	83	137411	43.123	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	140508	48.728	ug/l	86
77) Bromobenzene	11.201	156	142819	48.370	ug/l	78
78) n-propylbenzene	11.305	91	531715	43.456	ug/l	92
79) 2-Chlorotoluene	11.366	91	318668	43.103	ug/l	90
80) 1,3,5-Trimethylbenzene	11.451	105	422294	45.174	ug/l	94
81) trans-1,4-Dichloro-2-b...	11.018	75	31694	37.732	ug/l	97
82) 4-Chlorotoluene	11.457	91	367911	43.114	ug/l	90
83) tert-Butylbenzene	11.713	119	428947	45.398	ug/l	93
84) 1,2,4-Trimethylbenzene	11.750	105	420878	45.237	ug/l	94
85) sec-Butylbenzene	11.890	105	513700	45.162	ug/l	94
86) p-Isopropyltoluene	12.012	119	464131	46.756	ug/l	96
87) 1,3-Dichlorobenzene	11.969	146	266945	47.915	ug/l	98
88) 1,4-Dichlorobenzene	12.043	146	271660	46.686	ug/l	96
89) n-Butylbenzene	12.335	91	343330	43.788	ug/l	96
90) Hexachloroethane	12.536	117	66756	37.278	ug/l	76
91) 1,2-Dichlorobenzene	12.335	146	262072	46.971	ug/l	97
92) 1,2-Dibromo-3-Chloropr...	12.945	75	28501	42.534	ug/l	52
93) 1,2,4-Trichlorobenzene	13.591	180	179626	52.603	ug/l	97
94) Hexachlorobutadiene	13.725	225	73681	52.437	ug/l	99
95) Naphthalene	13.774	128	551611	51.859	ug/l	100
96) 1,2,3-Trichlorobenzene	13.963	180	178451	53.891	ug/l	98


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100422\
Data File : VX031880.D
Acq On : 04 Oct 2022 20:12
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Quant Time: Oct 05 06:28:20 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100322W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 04 06:59:30 2022
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed






#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX031880.D
Acq: 04 Oct 2022 20:12

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC







Tgt Ion:168 Resp: 139171
Ion Ratio Lower Upper
168 100
99 48.3 48.8 73.2#



#2
Dichlorodifluoromethane
Concen: 43.753 ug/l
RT: 1.167 min Scan# 13
Delta R.T. 0.001 min
Lab File: VX031880.D
Acq: 04 Oct 2022 20:12

Tgt Ion: 85 Resp: 78719
Ion Ratio Lower Upper
85 100
87 32.6 15.8 47.4









#3
Chloromethane
Concen: 33.324 ug/l
RT: 1.288 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX031880.D
Acq: 04 Oct 2022 20:12




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 50 Resp: 62851
Ion Ratio Lower Upper
50 100
52 35.0 26.9 40.3



#4
Vinyl Chloride
Concen: 35.135 ug/l
RT: 1.368 min Scan# 46
Delta R.T. -0.006 min
Lab File: VX031880.D
Acq: 04 Oct 2022 20:12

Tgt Ion: 62 Resp: 84413
Ion Ratio Lower Upper
62 100
64 32.9 25.3 37.9









#5
Bromomethane
Concen: 45.148 ug/l
RT: 1.612 min Scan# 80
Delta R.T. 0.001 min
Lab File: VX031880.D
Acq: 04 Oct 2022 20:12




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 94 Resp: 70551
Ion Ratio Lower Upper
94 100
96 95.4 74.4 111.6



#6
Chloroethane
Concen: 38.376 ug/l
RT: 1.685 min Scan# 98
Delta R.T. -0.000 min
Lab File: VX031880.D
Acq: 04 Oct 2022 20:12

Tgt Ion: 64 Resp: 79371
Ion Ratio Lower Upper
64 100
66 32.8 25.0 37.4









#7
Trichlorofluoromethane
Concen: 34.895 ug/l
RT: 1.886 min Scan# 11
Delta R.T. 0.006 min
Lab File: VX031880.D
Acq: 04 Oct 2022 20:12




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:101 Resp: 171090
Ion Ratio Lower Upper
101 100
103 63.3 51.8 77.6



#8
Diethyl Ether
Concen: 41.833 ug/l
RT: 2.130 min Scan# 171
Delta R.T. -0.000 min
Lab File: VX031880.D
Acq: 04 Oct 2022 20:12

Tgt Ion: 74 Resp: 55294
Ion Ratio Lower Upper
74 100
45 71.2 48.5 145.6









#9
1,1,2-Trichlorotrifluoroethane
Concen: 40.986 ug/l
RT: 2.325 min Scan# 204
Delta R.T. -0.000 min
Lab File: VX031880.D
Acq: 04 Oct 2022 20:12




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:101 Resp: 86811
Ion Ratio Lower Upper
101 100
85 42.7 38.6 57.8
151 89.4 54.0 81.0#



#10
Methyl Iodide
Concen: 51.945 ug/l
RT: 2.453 min Scan# 224
Delta R.T. -0.000 min
Lab File: VX031880.D
Acq: 04 Oct 2022 20:12

Tgt Ion:142 Resp: 103280
Ion Ratio Lower Upper
142 100
127 43.7 37.4 56.0
141 13.9 11.6 17.4









#11
Tert butyl alcohol
Concen: 209.410 ug/l
RT: 3.008 min Scan# 316
Delta R.T. -0.073 min
Lab File: VX031880.D
Acq: 04 Oct 2022 20:12




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 59 Resp: 72692
Ion Ratio Lower Upper
59 100
57 3.1 8.2 12.2#



#12
1,1-Dichloroethene
Concen: 41.960 ug/l
RT: 2.319 min Scan# 202
Delta R.T. -0.000 min
Lab File: VX031880.D
Acq: 04 Oct 2022 20:12

Tgt Ion: 96 Resp: 81608
Ion Ratio Lower Upper
96 100
61 124.7 135.3 202.9#
98 63.8 49.4 74.2





#13

Acrolein

Concen: 215.195 ug/l

RT: 2.240 min Scan# 11

Delta R.T. 0.001 min

Lab File: VX031880.D

Acq: 04 Oct 2022 20:12

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 56 Resp: 44226

Ion Ratio Lower Upper

56 100

55 68.8 56.0 84.0



#14

Allyl chloride

Concen: 34.871 ug/l

RT: 2.660 min Scan# 258

Delta R.T. 0.000 min

Lab File: VX031880.D

Acq: 04 Oct 2022 20:12




Tgt Ion: 41 Resp: 86570


Ion Ratio Lower Upper

41 100

39 73.2 58.2 87.4

76 48.6 26.9 40.3#









#15
Acrylonitrile
Concen: 197.840 ug/l
RT: 3.069 min Scan# 31
Delta R.T. 0.001 min
Lab File: VX031880.D
Acq: 04 Oct 2022 20:12




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 53 Resp: 179388
Ion Ratio Lower Upper
53 100
52 81.3 66.9 100.3
51 35.6 30.0 45.0



#16
Acetone
Concen: 229.472 ug/l
RT: 2.392 min Scan# 214
Delta R.T. -0.000 min
Lab File: VX031880.D
Acq: 04 Oct 2022 20:12

Tgt Ion: 43 Resp: 133524
Ion Ratio Lower Upper
43 100
58 38.1 23.8 35.8#





#17

Carbon Disulfide

Concen: 37.750 ug/l

RT: 2.508 min Scan# 21

Delta R.T. -0.000 min

Lab File: VX031880.D

Acq: 04 Oct 2022 20:12

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 76 Resp: 172479

Ion Ratio Lower Upper

76 100

78 9.7 7.2 10.8



#18

Methyl Acetate

Concen: 39.836 ug/l

RT: 2.703 min Scan# 265

Delta R.T. -0.006 min

Lab File: VX031880.D




Acq: 04 Oct 2022 20:12

Tgt Ion: 43 Resp: 92909

Ion Ratio Lower Upper

43 100

74 32.2 18.6 27.8#









#19
Methyl tert-butyl Ether
Concen: 40.793 ug/l
RT: 3.111 min Scan# 311
Delta R.T. 0.000 min
Lab File: VX031880.D
Acq: 04 Oct 2022 20:12




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 73 Resp: 258281
Ion Ratio Lower Upper
73 100
57 18.5 17.2 25.8



#20
Methylene Chloride
Concen: 40.745 ug/l
RT: 2.788 min Scan# 279
Delta R.T. 0.000 min
Lab File: VX031880.D
Acq: 04 Oct 2022 20:12

Tgt Ion: 84 Resp: 88874
Ion Ratio Lower Upper
84 100
49 92.5 105.9 158.9#
51 29.4 32.2 48.2#
86 65.2 50.8 76.2









#21
trans-1,2-Dichloroethene
Concen: 40.509 ug/l
RT: 3.093 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX031880.D
Acq: 04 Oct 2022 20:12




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 96 Resp: 92362
Ion Ratio Lower Upper
96 100
61 101.4 106.7 160.1#
98 63.6 48.7 73.1



#22
Diisopropyl ether
Concen: 37.774 ug/l
RT: 3.764 min Scan# 439
Delta R.T. 0.007 min
Lab File: VX031880.D
Acq: 04 Oct 2022 20:12

Tgt Ion: 45 Resp: 198844
Ion Ratio Lower Upper
45 100
43 68.6 59.0 88.6
87 42.6 22.6 33.8#
59 13.3 8.8 13.2#









#23
 Vinyl Acetate
 Concen: 197.697 ug/l
 RT: 3.721 min Scan# 41
 Delta R.T. -0.000 min
 Lab File: VX031880.D
 Acq: 04 Oct 2022 20:12




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion: 43 Resp: 778460
 Ion Ratio Lower Upper
 43 100
 86 16.7 8.7 13.1#



#24
 1,1-Dichloroethane
 Concen: 38.546 ug/l
 RT: 3.611 min Scan# 414
 Delta R.T. 0.000 min
 Lab File: VX031880.D
 Acq: 04 Oct 2022 20:12

Tgt Ion: 63 Resp: 136363
 Ion Ratio Lower Upper
 63 100
 98 9.2 3.5 10.4
 100 6.0 2.4 7.2









#25
2-Butanone
Concen: 192.051 ug/l
RT: 4.562 min Scan# 570
Delta R.T. -0.006 min
Lab File: VX031880.D
Acq: 04 Oct 2022 20:12




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 43 Resp: 219723
Ion Ratio Lower Upper
43 100
72 36.3 21.1 31.7#



#26
2,2-Dichloropropane
Concen: 33.570 ug/l
RT: 4.477 min Scan# 556
Delta R.T. -0.000 min
Lab File: VX031880.D
Acq: 04 Oct 2022 20:12

Tgt Ion: 77 Resp: 101388
Ion Ratio Lower Upper
77 100
97 26.8 10.9 32.6





#27

cis-1,2-Dichloroethene

Concen: 41.261 ug/l

RT: 4.489 min Scan# 558

Delta R.T. 0.000 min

Lab File: VX031880.D

Acq: 04 Oct 2022 20:12

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 96 Resp: 109385

Ion Ratio Lower Upper

96 100

61 108.8 0.0 305.2

98 65.7 0.0 128.4



#28

Bromochloromethane

Concen: 37.312 ug/l

RT: 4.898 min Scan# 625

Delta R.T. 0.001 min

Lab File: VX031880.D

Acq: 04 Oct 2022 20:12




Tgt Ion: 49 Resp: 53346


Ion Ratio Lower Upper

49 100

129 3.0 0.0 3.6

130 130.5 59.0 88.4#





#29

Tetrahydrofuran

Concen: 198.555 ug/l

RT: 5.013 min Scan# 643

Delta R.T. 0.000 min

Lab File: VX031880.D

Acq: 04 Oct 2022 20:12

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 42 Resp: 140260

Ion Ratio Lower Upper

42 100

72 63.3 36.4 54.6#

71 57.7 33.5 50.3#



#30

Chloroform

Concen: 39.075 ug/l

RT: 5.093 min Scan# 657

Delta R.T. 0.001 min

Lab File: VX031880.D




Acq: 04 Oct 2022 20:12


Tgt Ion: 83 Resp: 165371

Ion Ratio Lower Upper

83 100


85 65.0 52.8 79.2








#31
Cyclohexane
Concen: 37.986 ug/l
RT: 5.471 min Scan# 719
Delta R.T. 0.001 min
Lab File: VX031880.D
Acq: 04 Oct 2022 20:12

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC







Tgt Ion: 56 Resp: 112799
Ion Ratio Lower Upper
56 100
69 37.5 25.1 37.7
84 110.5 70.2 105.4#



#32
1,1,1-Trichloroethane
Concen: 38.276 ug/l
RT: 5.385 min Scan# 705
Delta R.T. 0.000 min
Lab File: VX031880.D
Acq: 04 Oct 2022 20:12

Tgt Ion: 97 Resp: 153720
Ion Ratio Lower Upper
97 100
99 65.9 51.3 76.9
61 35.1 35.9 53.9#





#33

1,2-Dichloroethane-d4

Concen: 37.070 ug/l

RT: 5.958 min Scan# 799

Delta R.T. 0.006 min

Lab File: VX031880.D

Acq: 04 Oct 2022 20:12

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 65 Resp: 92427

Ion Ratio Lower Upper

65 100

67 53.3 0.0 102.4



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.763 min Scan# 931

Delta R.T. 0.000 min

Lab File: VX031880.D

Acq: 04 Oct 2022 20:12




Tgt Ion: 114 Resp: 191077


Ion Ratio Lower Upper

114 100

63 14.6 0.0 42.6

88 14.8 0.0 33.0





#35

Dibromofluoromethane

Concen: 43.665 ug/l

RT: 5.385 min Scan# 705

Delta R.T. 0.000 min

Lab File: VX031880.D

Acq: 04 Oct 2022 20:12

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion:113 Resp: 99128

Ion Ratio Lower Upper

113 100

111 102.2 82.4 123.6

192 22.9 14.1 21.1#



#36

1,1-Dichloropropene

Concen: 40.362 ug/l

RT: 5.696 min Scan# 756

Delta R.T. 0.000 min

Lab File: VX031880.D

Acq: 04 Oct 2022 20:12




Tgt Ion: 75 Resp: 113617


Ion Ratio Lower Upper

75 100

110 44.8 17.5 52.5

77 31.9 24.9 37.3





#37

Ethyl Acetate

Concen: 40.957 ug/l

RT: 4.715 min Scan# 595

Delta R.T. 0.001 min

Lab File: VX031880.D

Acq: 04 Oct 2022 20:12

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 43 Resp: 87271

Ion Ratio Lower Upper

43 100

61 17.1 11.1 16.7#

70 15.4 8.6 13.0#



#38

Carbon Tetrachloride

Concen: 40.279 ug/l

RT: 5.678 min Scan# 753

Delta R.T. -0.000 min

Lab File: VX031880.D

Acq: 04 Oct 2022 20:12




Tgt Ion: 117 Resp: 140756


Ion Ratio Lower Upper

117 100

119 99.4 74.6 112.0

121 31.9 25.3 37.9





#39

Methylcyclohexane

Concen: 43.047 ug/l

RT: 7.379 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX031880.D

Acq: 04 Oct 2022 20:12

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 83 Resp: 148656

Ion Ratio Lower Upper

83 100

55 57.3 61.0 91.4#

98 53.4 37.7 56.5



#40

Benzene

Concen: 42.595 ug/l

RT: 6.038 min Scan# 812

Delta R.T. 0.001 min

Lab File: VX031880.D




Acq: 04 Oct 2022 20:12


Tgt Ion: 78 Resp: 346375

Ion Ratio Lower Upper

78 100

77 23.8 18.7 28.1









#41
Methacrylonitrile
Concen: 40.797 ug/l
RT: 4.922 min Scan# 61
Delta R.T. -0.000 min
Lab File: VX031880.D
Acq: 04 Oct 2022 20:12




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 41	Resp: 47238
Ion Ratio	Lower Upper
41 100	
39 54.4	45.0 67.6
67 95.1	58.2 87.4#
52 34.3	25.0 37.6



#42
1,2-Dichloroethane
Concen: 39.692 ug/l
RT: 6.086 min Scan# 820
Delta R.T. 0.000 min
Lab File: VX031880.D
Acq: 04 Oct 2022 20:12

Tgt Ion: 62	Resp: 113077
Ion Ratio	Lower Upper
62 100	
98 12.0	0.0 17.8









#43
Isopropyl Acetate
Concen: 41.881 ug/l
RT: 6.342 min Scan# 862
Delta R.T. 0.000 min
Lab File: VX031880.D
Acq: 04 Oct 2022 20:12




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 43 Resp: 144291
Ion Ratio Lower Upper
43 100
61 23.5 15.3 22.9#
87 19.8 10.6 15.8#



#44
Trichloroethene
Concen: 46.441 ug/l
RT: 7.129 min Scan# 991
Delta R.T. -0.000 min
Lab File: VX031880.D
Acq: 04 Oct 2022 20:12

Tgt Ion:130 Resp: 122007
Ion Ratio Lower Upper
130 100
95 84.5 0.0 206.8





#45

1,2-Dichloropropane

Concen: 40.512 ug/l

RT: 7.434 min Scan# 1041

Delta R.T. 0.007 min

Lab File: VX031880.D

Acq: 04 Oct 2022 20:12

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 63 Resp: 77180

Ion Ratio Lower Upper

63 100

65 32.5 26.6 39.8



#46

Dibromomethane

Concen: 43.198 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. 0.000 min

Lab File: VX031880.D

Acq: 04 Oct 2022 20:12




Tgt Ion: 93 Resp: 68596


Ion Ratio Lower Upper

93 100

95 83.8 65.5 98.3

174 125.4 67.6 101.4#









#47
Bromodichloromethane
Concen: 38.791 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. -0.000 min
Lab File: VX031880.D
Acq: 04 Oct 2022 20:12




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 83 Resp: 123280
Ion Ratio Lower Upper
83 100
85 65.7 51.0 76.6
127 10.3 7.0 10.6



#48
Methyl methacrylate
Concen: 43.258 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. -0.000 min
Lab File: VX031880.D
Acq: 04 Oct 2022 20:12

Tgt Ion: 41 Resp: 66692
Ion Ratio Lower Upper
41 100
69 113.0 64.6 96.8#
39 65.1 51.4 77.0









#49
1,4-Dioxane
Concen: 867.218 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. -0.018 min
Lab File: VX031880.D
Acq: 04 Oct 2022 20:12




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 88 Resp: 43335
Ion Ratio Lower Upper
88 100
43 22.6 27.1 40.7#
58 53.4 53.9 80.9#



#50
Toluene-d8
Concen: 42.942 ug/l
RT: 8.653 min Scan# 1241
Delta R.T. -0.000 min
Lab File: VX031880.D
Acq: 04 Oct 2022 20:12

Tgt Ion: 98 Resp: 348746
Ion Ratio Lower Upper
98 100
100 67.6 52.5 78.7





#51

4-Methyl-2-Pentanone

Concen: 210.448 ug/l

RT: 8.574 min Scan# 11

Delta R.T. -0.006 min

Lab File: VX031880.D

Acq: 04 Oct 2022 20:12

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 43 Resp: 455122

Ion Ratio Lower Upper

43 100

58 45.6 30.6 46.0



#52

Toluene

Concen: 42.720 ug/l

RT: 8.720 min Scan# 1252

Delta R.T. 0.000 min

Lab File: VX031880.D




Acq: 04 Oct 2022 20:12


Tgt Ion: 92 Resp: 243547

Ion Ratio Lower Upper

92 100

91 168.4 137.1 205.7





#53

t-1,3-Dichloropropene

Concen: 39.758 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX031880.D

Acq: 04 Oct 2022 20:12

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 75 Resp: 120485

Ion Ratio Lower Upper

75 100

77 31.0 24.8 37.2



#54

cis-1,3-Dichloropropene

Concen: 40.649 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX031880.D

Acq: 04 Oct 2022 20:12




Tgt Ion: 75 Resp: 134983


Ion Ratio Lower Upper

75 100

77 30.3 24.4 36.6

39 37.1 43.7 65.5#









#55
 1,1,2-Trichloroethane
 Concen: 44.113 ug/l
 RT: 9.153 min Scan# 11
 Delta R.T. -0.000 min
 Lab File: VX031880.D
 Acq: 04 Oct 2022 20:12

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC

Tgt Ion: 97 Resp: 100238




Ion	Ratio	Lower	Upper
97	100		
83	78.4	69.0	103.6
85	51.7	44.1	66.1
99	63.0	51.3	76.9




#56
 Ethyl methacrylate
 Concen: 48.209 ug/l
 RT: 9.116 min Scan# 1317
 Delta R.T. 0.000 min
 Lab File: VX031880.D
 Acq: 04 Oct 2022 20:12

Tgt Ion: 69 Resp: 126664

Ion	Ratio	Lower	Upper
69	100		
41	50.2	54.2	81.4#
39	32.5	34.2	51.2#





#57

1,3-Dichloropropane

Concen: 42.407 ug/l

RT: 9.311 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX031880.D

Acq: 04 Oct 2022 20:12

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 76 Resp: 149650

Ion Ratio Lower Upper

76 100

78 33.3 25.5 38.3



#58

2-Chloroethyl Vinyl ether

Concen: 223.373 ug/l

RT: 8.245 min Scan# 1174

Delta R.T. 0.001 min

Lab File: VX031880.D




Acq: 04 Oct 2022 20:12


Tgt Ion: 63 Resp: 302427

Ion Ratio Lower Upper

63 100

106 39.4 22.6 34.0#





#59

2-Hexanone

Concen: 216.015 ug/l

RT: 9.433 min Scan# 1369

Delta R.T. 0.000 min

Lab File: VX031880.D

Acq: 04 Oct 2022 20:12

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 43 Resp: 339620

Ion Ratio Lower Upper

43 100

58 63.7 26.7 80.0



#60

Dibromochloromethane

Concen: 41.653 ug/l

RT: 9.525 min Scan# 1384

Delta R.T. -0.000 min

Lab File: VX031880.D




Acq: 04 Oct 2022 20:12


Tgt Ion: 129 Resp: 114902

Ion Ratio Lower Upper

129 100

127 76.0 38.7 116.1









#61
1,2-Dibromoethane
Concen: 44.215 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. 0.000 min
Lab File: VX031880.D
Acq: 04 Oct 2022 20:12




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:107 Resp: 111566
Ion Ratio Lower Upper
107 100
109 95.3 74.2 111.4



#62
4-Bromofluorobenzene
Concen: 43.659 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX031880.D
Acq: 04 Oct 2022 20:12

Tgt Ion: 95 Resp: 127682
Ion Ratio Lower Upper
95 100
174 101.8 0.0 153.2
176 99.3 0.0 144.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX031880.D
Acq: 04 Oct 2022 20:12




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:117 Resp: 206110
Ion Ratio Lower Upper
117 100
82 46.1 46.8 70.2#
119 32.7 26.4 39.6



#64
Tetrachloroethene
Concen: 49.268 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.000 min
Lab File: VX031880.D
Acq: 04 Oct 2022 20:12

Tgt Ion:164 Resp: 126951
Ion Ratio Lower Upper
164 100
166 129.4 99.0 148.4
129 85.2 78.0 117.0
131 82.6 74.2 111.2









#65
Chlorobenzene
Concen: 45.841 ug/l
RT: 10.080 min Scan# 1475
Delta R.T. 0.001 min
Lab File: VX031880.D
Acq: 04 Oct 2022 20:12




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:112 Resp: 286485
Ion Ratio Lower Upper
112 100
114 33.0 25.7 38.5



#66
1,1,1,2-Tetrachloroethane
Concen: 45.049 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. -0.000 min
Lab File: VX031880.D
Acq: 04 Oct 2022 20:12

Tgt Ion:131 Resp: 108365
Ion Ratio Lower Upper
131 100
133 97.1 46.6 139.7
119 62.1 31.9 95.9





#67

Ethyl Benzene

Concen: 45.208 ug/l

RT: 10.195 min Scan# 1494

Delta R.T. 0.000 min

Lab File: VX031880.D

Acq: 04 Oct 2022 20:12

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 91 Resp: 465068

Ion Ratio Lower Upper

91 100

106 34.4 25.2 37.8



#68

m/p-Xylenes

Concen: 93.266 ug/l

RT: 10.299 min Scan# 1511

Delta R.T. -0.006 min

Lab File: VX031880.D




Acq: 04 Oct 2022 20:12


Tgt Ion: 106 Resp: 396107

Ion Ratio Lower Upper

106 100

91 182.0 161.0 241.4









#69
o-Xylene
Concen: 48.032 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. 0.000 min
Lab File: VX031880.D
Acq: 04 Oct 2022 20:12




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:106 Resp: 196530
Ion Ratio Lower Upper
106 100
91 192.8 107.2 321.6



#70
Styrene
Concen: 48.072 ug/l
RT: 10.659 min Scan# 1570
Delta R.T. -0.000 min
Lab File: VX031880.D
Acq: 04 Oct 2022 20:12

Tgt Ion:104 Resp: 319702
Ion Ratio Lower Upper
104 100
78 43.1 41.4 62.2
103 53.2 43.8 65.8









#71
Bromoform
Concen: 42.961 ug/l
RT: 10.799 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX031880.D
Acq: 04 Oct 2022 20:12




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:173 Resp: 86905
Ion Ratio Lower Upper
173 100
175 48.0 24.0 72.0
254 0.1 0.0 0.0#



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX031880.D
Acq: 04 Oct 2022 20:12

Tgt Ion:152 Resp: 158486
Ion Ratio Lower Upper
152 100
115 74.3 43.4 130.1
150 174.5 0.0 346.8









#73
Isopropylbenzene
Concen: 44.665 ug/l
RT: 10.964 min Scan# 1620
Delta R.T. 0.001 min
Lab File: VX031880.D
Acq: 04 Oct 2022 20:12




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:105 Resp: 500006
Ion Ratio Lower Upper
105 100
120 28.9 13.7 41.0



#74
N-ethyl acetate
Concen: 46.625 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. -0.006 min
Lab File: VX031880.D
Acq: 04 Oct 2022 20:12

Tgt Ion: 43 Resp: 122246
Ion Ratio Lower Upper
43 100
70 56.5 33.4 50.2#
55 27.2 18.2 27.2
61 28.6 19.3 28.9









#75
 1,1,2,2-Tetrachloroethane
 Concen: 43.123 ug/l
 RT: 11.214 min Scan# 1661
 Delta R.T. 0.001 min
 Lab File: VX031880.D
 Acq: 04 Oct 2022 20:12




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion: 83 Resp: 137411
 Ion Ratio Lower Upper
 83 100
 131 13.1 5.1 15.4
 85 65.5 32.6 97.8



#76
 1,2,3-Trichloropropane
 Concen: 48.728 ug/l
 RT: 11.238 min Scan# 1665
 Delta R.T. -0.006 min
 Lab File: VX031880.D
 Acq: 04 Oct 2022 20:12

Tgt Ion: 75 Resp: 140508
 Ion Ratio Lower Upper
 75 100
 77 32.0 20.3 60.9





#77

Bromobenzene

Concen: 48.370 ug/l

RT: 11.201 min Scan# 1659

Delta R.T. 0.000 min

Lab File: VX031880.D

Acq: 04 Oct 2022 20:12

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion:156 Resp: 142819

Ion Ratio Lower Upper

156 100

77 112.8 79.5 238.6

158 97.3 48.4 145.3



#78

n-propylbenzene

Concen: 43.456 ug/l

RT: 11.305 min Scan# 1676

Delta R.T. -0.000 min

Lab File: VX031880.D




Acq: 04 Oct 2022 20:12


Tgt Ion: 91 Resp: 531715

Ion Ratio Lower Upper

91 100

120 27.5 11.8 35.4









#79
2-Chlorotoluene
Concen: 43.103 ug/l
RT: 11.366 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX031880.D
Acq: 04 Oct 2022 20:12




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 91 Resp: 318668
Ion Ratio Lower Upper
91 100
126 40.4 17.4 52.3



#80
1,3,5-Trimethylbenzene
Concen: 45.174 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX031880.D
Acq: 04 Oct 2022 20:12

Tgt Ion: 105 Resp: 422294
Ion Ratio Lower Upper
105 100
120 52.9 24.5 73.5









#81
trans-1,4-Dichloro-2-butene
Concen: 37.732 ug/l
RT: 11.018 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX031880.D
Acq: 04 Oct 2022 20:12




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 75 Resp: 31694
Ion Ratio Lower Upper
75 100
53 96.0 77.6 116.4
89 50.7 37.4 56.0



#82
4-Chlorotoluene
Concen: 43.114 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VX031880.D
Acq: 04 Oct 2022 20:12

Tgt Ion: 91 Resp: 367911
Ion Ratio Lower Upper
91 100
126 35.4 15.1 45.3









#83
tert-Butylbenzene
Concen: 45.398 ug/l
RT: 11.713 min Scan# 1743
Delta R.T. 0.000 min
Lab File: VX031880.D
Acq: 04 Oct 2022 20:12




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:119 Resp: 428947
Ion Ratio Lower Upper
119 100
91 52.4 29.2 87.6
134 26.0 11.6 34.8



#84
1,2,4-Trimethylbenzene
Concen: 45.237 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. -0.000 min
Lab File: VX031880.D
Acq: 04 Oct 2022 20:12

Tgt Ion:105 Resp: 420878
Ion Ratio Lower Upper
105 100
120 49.0 22.6 67.8









#85
 sec-Butylbenzene
 Concen: 45.162 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. 0.000 min
 Lab File: VX031880.D
 Acq: 04 Oct 2022 20:12




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion:105 Resp: 513700
 Ion Ratio Lower Upper
 105 100
 134 23.3 10.4 31.1



#86
 p-Isopropyltoluene
 Concen: 46.756 ug/l
 RT: 12.012 min Scan# 1792
 Delta R.T. 0.000 min
 Lab File: VX031880.D
 Acq: 04 Oct 2022 20:12

Tgt Ion:119 Resp: 464131
 Ion Ratio Lower Upper
 119 100
 134 27.8 13.4 40.2
 91 21.2 12.4 37.2





#87

1,3-Dichlorobenzene

Concen: 47.915 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. 0.000 min

Lab File: VX031880.D

Acq: 04 Oct 2022 20:12

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion:146 Resp: 266945

Ion Ratio Lower Upper

146 100

111 35.8 19.8 59.4

148 63.0 31.5 94.5



#88

1,4-Dichlorobenzene

Concen: 46.686 ug/l

RT: 12.043 min Scan# 1797

Delta R.T. 0.001 min

Lab File: VX031880.D

Acq: 04 Oct 2022 20:12




Tgt Ion:146 Resp: 271660


Ion Ratio Lower Upper

146 100

111 34.8 20.5 61.5

148 63.4 32.0 96.2









#89
n-Butylbenzene
Concen: 43.788 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX031880.D
Acq: 04 Oct 2022 20:12




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 91 Resp: 343330
Ion Ratio Lower Upper
91 100
92 54.8 27.0 81.0
134 32.1 13.3 39.9



#90
Hexachloroethane
Concen: 37.278 ug/l
RT: 12.536 min Scan# 1878
Delta R.T. -0.006 min
Lab File: VX031880.D
Acq: 04 Oct 2022 20:12

Tgt Ion: 117 Resp: 66756
Ion Ratio Lower Upper
117 100
201 89.4 34.8 104.4





#91

1,2-Dichlorobenzene

Concen: 46.971 ug/l

RT: 12.335 min Scan# 1845

Delta R.T. 0.000 min

Lab File: VX031880.D

Acq: 04 Oct 2022 20:12

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion:146 Resp: 262072

Ion Ratio Lower Upper

146 100

111 37.1 20.6 61.8

148 64.4 31.6 95.0



#92

1,2-Dibromo-3-Chloropropane

Concen: 42.534 ug/l

RT: 12.945 min Scan# 1945

Delta R.T. -0.000 min

Lab File: VX031880.D

Acq: 04 Oct 2022 20:12




Tgt Ion: 75 Resp: 28501


Ion Ratio Lower Upper

75 100

155 119.6 40.3 120.8

157 154.4 51.5 154.5









#93
 1,2,4-Trichlorobenzene
 Concen: 52.603 ug/l
 RT: 13.591 min Scan# 2051
 Delta R.T. 0.000 min
 Lab File: VX031880.D
 Acq: 04 Oct 2022 20:12




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion:180 Resp: 179626
 Ion Ratio Lower Upper
 180 100
 182 94.6 46.9 140.6
 145 29.1 16.7 50.1



#94
 Hexachlorobutadiene
 Concen: 52.437 ug/l
 RT: 13.725 min Scan# 2073
 Delta R.T. 0.000 min
 Lab File: VX031880.D
 Acq: 04 Oct 2022 20:12

Tgt Ion:225 Resp: 73681
 Ion Ratio Lower Upper
 225 100
 223 62.1 30.8 92.4
 227 64.0 32.3 96.9









#95
Naphthalene
Concen: 51.859 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. 0.000 min
Lab File: VX031880.D
Acq: 04 Oct 2022 20:12

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tgt Ion:128 Resp: 551611
Ion Ratio Lower Upper
128 100
127 12.9 10.4 15.6
129 11.1 8.8 13.2



#96
1,2,3-Trichlorobenzene
Concen: 53.891 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. 0.000 min
Lab File: VX031880.D
Acq: 04 Oct 2022 20:12

Tgt Ion:180 Resp: 178451
Ion Ratio Lower Upper
180 100
182 95.5 47.9 143.8
145 31.3 17.6 52.8

