

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100422\
 Data File : VX031867.D
 Acq On : 04 Oct 2022 14:24
 Operator : JC/MD
 Sample : N4914-04DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S2DL

Quant Time: Oct 05 06:20:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100322W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 04 06:59:30 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	103672	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	152504	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	171938	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	122944	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	85719	46.151	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	92.300%
35) Dibromofluoromethane	5.385	113	82365	45.457	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.920%
50) Toluene-d8	8.647	98	310657	47.927	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.860%
62) 4-Bromofluorobenzene	11.079	95	114104	48.885	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.760%
Target Compounds						
					Qvalue	
40) Benzene	6.038	78	250045	38.526	ug/l	99
43) Isopropyl Acetate	6.342	43	10746	3.908	ug/l #	91
52) Toluene	8.720	92	138623	30.466	ug/l	98
67) Ethyl Benzene	10.195	91	68948	8.034	ug/l	95
68) m/p-Xylenes	10.299	106	57374	16.194	ug/l	92
69) o-Xylene	10.640	106	37984	11.128	ug/l	93
73) Isopropylbenzene	10.963	105	4228	0.487	ug/l	95
80) 1,3,5-Trimethylbenzene	11.451	105	11330	1.562	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	29153	4.039	ug/l	92
95) Naphthalene	13.780	128	4421909	535.897	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100422\
Data File : VX031867.D
Acq On : 04 Oct 2022 14:24
Operator : JC/MD
Sample : N4914-04DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S2DL

Quant Time: Oct 05 06:20:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100322W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 04 06:59:30 2022
Response via : Initial Calibration

