

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100423\
 Data File : VX038060.D
 Acq On : 04 Oct 2023 13:15
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Oct 05 01:01:18 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100323W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 04 09:08:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	108907	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	193796	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	186687	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	94754	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	81292	47.973	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.940%
35) Dibromofluoromethane	5.385	113	62410	46.942	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.880%
50) Toluene-d8	8.652	98	243162	47.791	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.580%
62) 4-Bromofluorobenzene	11.079	95	89305	48.227	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.460%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100423\
Data File : VX038060.D
Acq On : 04 Oct 2023 13:15
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Oct 05 01:01:18 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100323W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 04 09:08:35 2023
Response via : Initial Calibration

