

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100423\
 Data File : VX038064.D
 Acq On : 04 Oct 2023 14:48
 Operator : JC/MD
 Sample : 04656-25
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-10022023

Quant Time: Oct 05 01:02:37 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100323W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 04 09:08:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	98709	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	175232	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	171519	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	83621	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	73617	47.932	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.860%
35) Dibromofluoromethane	5.391	113	56756	47.212	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.420%
50) Toluene-d8	8.653	98	222611	48.387	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.780%
62) 4-Bromofluorobenzene	11.079	95	77809	46.470	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.940%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100423\
Data File : VX038064.D
Acq On : 04 Oct 2023 14:48
Operator : JC/MD
Sample : 04656-25
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-10022023

Quant Time: Oct 05 01:02:37 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100323W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 04 09:08:35 2023
Response via : Initial Calibration

