

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100423\
 Data File : VX038065.D
 Acq On : 04 Oct 2023 15:11
 Operator : JC/MD
 Sample : 04656-18
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-101S

Quant Time: Oct 05 01:02:54 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100323W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 04 09:08:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	101773	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	181744	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	180359	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	92654	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	76729	48.454	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.900%
35) Dibromofluoromethane	5.391	113	59288	47.551	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.100%
50) Toluene-d8	8.653	98	233262	48.885	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.760%
62) 4-Bromofluorobenzene	11.079	95	85756	49.381	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.760%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	1517	2.181	ug/l	92
64) Tetrachloroethene	9.275	164	2014	1.678	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100423\
Data File : VX038065.D
Acq On : 04 Oct 2023 15:11
Operator : JC/MD
Sample : 04656-18
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-101S

Quant Time: Oct 05 01:02:54 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100323W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 04 09:08:35 2023
Response via : Initial Calibration

