

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100423\
 Data File : VX038068.D
 Acq On : 04 Oct 2023 16:21
 Operator : JC/MD
 Sample : 04656-21
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-102S

Quant Time: Oct 05 01:03:55 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100323W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 04 09:08:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	97114	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	173082	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	170440	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	84912	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	74075	49.022	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.040%
35) Dibromofluoromethane	5.385	113	56025	47.182	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.360%
50) Toluene-d8	8.647	98	220845	48.599	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.200%
62) 4-Bromofluorobenzene	11.079	95	79931	48.331	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.660%
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.319	96	2213	2.407	ug/l	91
16) Acetone	2.386	43	1517	2.286	ug/l #	86
24) 1,1-Dichloroethane	3.617	63	3614	1.748	ug/l	98
27) cis-1,2-Dichloroethene	4.501	96	4892	3.798	ug/l	86
30) Chloroform	5.099	83	1670	0.742	ug/l	96
44) Trichloroethene	7.129	130	21248	15.950	ug/l	98
64) Tetrachloroethene	9.275	164	6640	5.855	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100423\
Data File : VX038068.D
Acq On : 04 Oct 2023 16:21
Operator : JC/MD
Sample : 04656-21
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-102S

Quant Time: Oct 05 01:03:55 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100323W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 04 09:08:35 2023
Response via : Initial Calibration

