

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100423\
 Data File : VX038069.D
 Acq On : 04 Oct 2023 16:45
 Operator : JC/MD
 Sample : 04656-22
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-1021

Quant Time: Oct 05 01:04:21 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100323W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 04 09:08:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	97359	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	174495	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	173006	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	90132	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	72659	47.964	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.920%
35) Dibromofluoromethane	5.385	113	56374	47.092	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.180%
50) Toluene-d8	8.653	98	222286	48.520	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.040%
62) 4-Bromofluorobenzene	11.079	95	81926	49.136	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.280%
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.319	96	9628	10.445	ug/l	91
24) 1,1-Dichloroethane	3.611	63	10071	4.860	ug/l	98
27) cis-1,2-Dichloroethene	4.489	96	13508	10.460	ug/l	82
30) Chloroform	5.099	83	4054	1.796	ug/l	99
44) Trichloroethene	7.129	130	65326	48.641	ug/l	99
64) Tetrachloroethene	9.275	164	21141	18.366	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100423\
Data File : VX038069.D
Acq On : 04 Oct 2023 16:45
Operator : JC/MD
Sample : 04656-22
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-102I

Quant Time: Oct 05 01:04:21 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100323W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 04 09:08:35 2023
Response via : Initial Calibration

