

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100423\
 Data File : VX038075.D
 Acq On : 04 Oct 2023 19:05
 Operator : JC/MD
 Sample : 04682-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-4

Quant Time: Oct 05 01:06:27 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100323W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 04 09:08:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	94475	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	170557	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	169712	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	85650	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	71586	48.698	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.400%
35) Dibromofluoromethane	5.391	113	54735	46.778	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.560%
50) Toluene-d8	8.653	98	216246	48.291	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.580%
62) 4-Bromofluorobenzene	11.079	95	78304	48.048	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.100%
Target Compounds						
16) Acetone	2.373	43	1674	2.593	ug/l	Qvalue # 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100423\
Data File : VX038075.D
Acq On : 04 Oct 2023 19:05
Operator : JC/MD
Sample : 04682-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-4

Quant Time: Oct 05 01:06:27 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100323W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 04 09:08:35 2023
Response via : Initial Calibration

